

Clear table


Instrument ID: IC-2 Analyst : IZ Method: 300.0 / 9056A

ident	concentratio tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio n HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-091624	2024-09-16 11:39:1	10	NF/IZ
STD2	0.424	0.637	0.645	2.133	0.536	1.071	3.306	IC1-091624	2024-09-16 12:00:1	10	NF/IZ
STD3	0.802	1.202	1.205	4.008	1.002	2.024	6.041	IC1-091624	2024-09-16 12:22:1	10	NF/IZ
STD4	1.007	1.523	1.517	5.064	1.267	2.521	7.591	IC1-091624	2024-09-16 12:43:1	10	NF/IZ
STD5	1.964	2.936	2.928	9.781	2.445	4.86	14.552	IC1-091624	2024-09-16 13:05:1	10	NF/IZ
STD6	3.962	5.921	5.921	19.755	4.924	9.91	29.354	IC1-091624	2024-09-16 13:26:1	10	NF/IZ
STD7	5.041	7.581	7.584	25.259	6.326	12.614	37.655	IC1-091624	2024-09-16 13:47:1	10	NF/IZ

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	6.0000	6.1667	7.5000	6.6500	7.2000	7.1000	10.2000
STD3	0.2500	0.1667	0.4167	0.2000	0.2000	1.2000	0.6833
STD4	0.7000	1.5333	1.1333	1.2800	1.3600	0.8400	1.2133
STD5	-1.8000	-2.1333	-2.4000	-2.1900	-2.2000	-2.8000	-2.9867
STD6	-0.9500	-1.3167	-1.3167	-1.2250	-1.5200	-0.9000	-2.1533
STD7	0.8200	1.0800	1.1200	1.0360	1.2160	0.9120	1.7703

Fluoride (Anions)




Function: $A = -5.20226E-3 + 0.0147183 \times Q$

Relative standard deviation 1.655264 %

Correlation coefficient 0.999855

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	0.400	10.0	1.0	1.0	0.057	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	0.800	10.0	1.0	1.0	0.113	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	1.000	10.0	1.0	1.0	0.143	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	2.000	10.0	1.0	1.0	0.284	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	4.000	10.0	1.0	1.0	0.578	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	5.000	10.0	1.0	1.0	0.737	STD7	2024-09-16 13:47:53 UTC-4	used


Chloride (Anions)



Function: $A = -4.73518E-3 + 9.92061E-3 \times Q$

Relative standard deviation 2.104599 %
Correlation coefficient 0.999766


Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.058	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.114	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.146	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.287	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	0.583	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	0.747	STD7	2024-09-16 13:47:53 UTC-4	used

Nitrite (Anions) $(\mu\text{S/cm}) \times \text{min}$ 

Function: $A = -0.0172576 + 0.0227000 \times Q$
Relative standard deviation 2.245267 %
Correlation coefficient 0.999738

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.129	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.256	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.327	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.647	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	1.327	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	1.704	STD7	2024-09-16 13:47:53 UTC-4	used

Bromide (Anions)




Function: $A = -8.29886E-3 + 4.43473E-3 \times Q$

Relative standard deviation 2.051807 %

Correlation coefficient 0.999778

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	2.000	10.0	1.0	1.0	0.086	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	4.000	10.0	1.0	1.0	0.169	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	5.000	10.0	1.0	1.0	0.216	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	10.000	10.0	1.0	1.0	0.425	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	20.000	10.0	1.0	1.0	0.868	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	25.000	10.0	1.0	1.0	1.112	STD7	2024-09-16 13:47:53 UTC-4	used

Nitrate (Anions)




Function: $A = -0.0140308 + 0.0249539 \times Q$

Relative standard deviation 2.352995 %

Correlation coefficient 0.999711

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	0.500	10.0	1.0	1.0	0.120	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	1.000	10.0	1.0	1.0	0.236	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	1.250	10.0	1.0	1.0	0.302	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	2.500	10.0	1.0	1.0	0.596	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	5.000	10.0	1.0	1.0	1.215	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	6.250	10.0	1.0	1.0	1.565	STD7	2024-09-16 13:47:53 UTC-4	used


Phosphate (Anions) $(\mu\text{S/cm}) \times \text{min}$ Function: $A = -0.0104137 + 8.60889E-3 \times Q$

Relative standard deviation 2.003622 %

Correlation coefficient 0.999791

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	1.000	10.0	1.0	1.0	0.082	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	2.000	10.0	1.0	1.0	0.164	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	2.500	10.0	1.0	1.0	0.207	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	5.000	10.0	1.0	1.0	0.408	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	10.000	10.0	1.0	1.0	0.843	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	12.500	10.0	1.0	1.0	1.075	STD7	2024-09-16 13:47:53 UTC-4	used

Sulfate (Anions)

 $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -0.0328706 + 7.71866E-3 \times Q$

Relative standard deviation 3.353471 %

Correlation coefficient 0.999413

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2024-09-16 11:39:22 UTC-4	used
Standard 2	1	3.000	10.0	1.0	1.0	0.222	STD2	2024-09-16 12:00:45 UTC-4	used
Standard 3	1	6.000	10.0	1.0	1.0	0.433	STD3	2024-09-16 12:22:09 UTC-4	used
Standard 4	1	7.500	10.0	1.0	1.0	0.553	STD4	2024-09-16 12:43:34 UTC-4	used
Standard 5	1	15.000	10.0	1.0	1.0	1.090	STD5	2024-09-16 13:05:00 UTC-4	used
Standard 6	1	30.000	10.0	1.0	1.0	2.233	STD6	2024-09-16 13:26:26 UTC-4	used
Standard 7	1	37.000	10.0	1.0	1.0	2.874	STD7	2024-09-16 13:47:53 UTC-4	used


Sample data

Ident STD1
Sample type Standard 1
Determination start 2024-09-16 11:39:22 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.39 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident STD2
Sample type Standard 2
Determination start 2024-09-16 12:00:45 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.33 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.428	0.0572	0.413	0.424	Fluoride
2	6.743	0.0585	0.322	0.637	Chloride
3	8.940	0.1292	0.555	0.645	Nitrite
4	11.067	0.0863	0.317	2.133	Bromide
5	12.818	0.1197	0.385	0.536	Nitrate
6	15.352	0.0818	0.171	1.071	Phosphate
7	16.930	0.2223	0.508	3.306	Sulfate


Sample data

Ident STD3
Sample type Standard 3
Determination start 2024-09-16 12:22:09 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.22 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.428	0.1129	0.820	0.802	Fluoride
2	6.735	0.1145	0.634	1.202	Chloride
3	8.927	0.2562	1.109	1.205	Nitrite
4	11.045	0.1695	0.629	4.008	Bromide
5	12.795	0.2360	0.764	1.002	Nitrate
6	15.353	0.1638	0.357	2.024	Phosphate
7	16.947	0.4334	1.003	6.041	Sulfate


Sample data

Ident STD4
Sample type Standard 4
Determination start 2024-09-16 12:43:34 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.11 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.433	0.1429	1.040	1.007	Fluoride
2	6.732	0.1464	0.811	1.523	Chloride
3	8.917	0.3271	1.418	1.517	Nitrite
4	11.027	0.2163	0.804	5.064	Bromide
5	12.772	0.3021	0.981	1.267	Nitrate
6	15.345	0.2066	0.448	2.521	Phosphate
7	16.958	0.5530	1.283	7.591	Sulfate


Sample data

Ident STD5
Sample type Standard 5
Determination start 2024-09-16 13:05:00 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.05 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.427	0.2838	2.088	1.964	Fluoride
2	6.718	0.2865	1.601	2.936	Chloride
3	8.897	0.6473	2.828	2.928	Nitrite
4	10.998	0.4255	1.596	9.781	Bromide
5	12.740	0.5961	1.949	2.445	Nitrate
6	15.323	0.4080	0.897	4.860	Phosphate
7	16.945	1.0904	2.544	14.552	Sulfate


Sample data

Ident STD6
Sample type Standard 6
Determination start 2024-09-16 13:26:26 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.22 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.423	0.5779	4.306	3.962	Fluoride
2	6.702	0.5827	3.302	5.921	Chloride
3	8.878	1.3268	5.858	5.921	Nitrite
4	10.965	0.8678	3.292	19.755	Bromide
5	12.698	1.2148	4.037	4.924	Nitrate
6	15.287	0.8428	1.914	9.910	Phosphate
7	16.922	2.2329	5.279	29.354	Sulfate


Sample data

Ident STD7
Sample type Standard 7
Determination start 2024-09-16 13:47:53 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.16 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.435	0.7368	5.489	5.041	Fluoride
2	6.715	0.7473	4.234	7.581	Chloride
3	8.892	1.7044	7.521	7.584	Nitrite
4	10.973	1.1119	4.227	25.259	Bromide
5	12.708	1.5645	5.205	6.326	Nitrate
6	15.313	1.0755	2.439	12.614	Phosphate
7	16.965	2.8736	6.798	37.655	Sulfate


Sample data

Ident ICV
Sample type Check standard 1
Determination start 2024-09-16 15:00:21 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 11.88 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.427	0.2818	2.081	1.950	Fluoride
2	6.685	0.2862	1.608	2.933	Chloride
3	8.842	0.6478	2.847	2.930	Nitrite
4	10.915	0.4262	1.608	9.798	Bromide
5	12.650	0.5947	1.960	2.440	Nitrate
6	15.303	0.4147	0.925	4.938	Phosphate
7	16.998	1.0857	2.548	14.491	Sulfate


Sample data

Ident ICB
Sample type Sample
Determination start 2024-09-16 15:21:50 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.22 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident CCV
Sample type Check standard 1
Determination start 2024-10-04 10:12:35 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.78 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.427	0.2927	1.910	2.024	Fluoride
2	6.802	0.2930	1.391	3.001	Chloride
3	9.052	0.6604	2.425	2.985	Nitrite
4	11.243	0.4373	1.353	10.047	Bromide
5	13.013	0.6092	1.645	2.497	Nitrate
6	14.792	0.4259	0.883	5.068	Phosphate
7	16.878	1.1061	2.262	14.756	Sulfate


Sample data

Ident CCB
Sample type Sample
Determination start 2024-10-04 10:34:05 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.73 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident LB132760BLW
Sample type Sample
Determination start 2024-10-04 10:55:36 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.61 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident LB132760BSW
Sample type Check standard 1
Determination start 2024-10-04 11:17:06 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.50 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.427	0.2934	1.950	2.028	Fluoride
2	6.763	0.2927	1.417	2.999	Chloride
3	8.983	0.6643	2.488	3.002	Nitrite
4	11.133	0.4372	1.383	10.046	Bromide
5	12.892	0.6089	1.679	2.496	Nitrate
6	14.782	0.4296	0.904	5.112	Phosphate
7	16.922	1.1072	2.291	14.771	Sulfate


Sample data

Ident P4285-03
Sample type Sample
Determination start 2024-10-04 11:38:39 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.50 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.430	0.0597	0.377	0.441	Fluoride
2	6.877	48.2715	207.784	486.626	Chloride
3	8.992	0.0208	0.082	0.168	Nitrite
4	11.133	0.0193	0.062	0.622	Bromide
5	12.880	1.0130	2.878	4.116	Nitrate
6	16.940	3.5012	7.421	45.786	Sulfate


Sample data

Ident P4285-03MS
Sample type Sample
Determination start 2024-10-04 12:00:12 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.67 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.427	0.3629	2.399	2.501	Fluoride
2	6.858	46.3906	203.088	467.667	Chloride
3	8.978	0.6956	2.722	3.140	Nitrite
4	11.100	0.4640	1.516	10.651	Bromide
5	12.847	1.6271	4.676	6.577	Nitrate
6	14.778	0.4440	0.936	5.279	Phosphate
7	16.933	4.5767	9.784	59.720	Sulfate


Sample data

Ident P4285-03MSD
Sample type Sample
Determination start 2024-10-04 12:21:45 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.33 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.423	0.3539	2.346	2.440	Fluoride
2	6.850	46.4856	204.168	468.624	Chloride
3	8.965	0.6719	2.641	3.036	Nitrite
4	11.080	0.4473	1.471	10.273	Bromide
5	12.825	1.6076	4.642	6.499	Nitrate
6	14.777	0.4278	0.901	5.090	Phosphate
7	16.942	4.5443	9.767	59.300	Sulfate


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2024-10-04 12:43:20 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.27 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.433	0.2881	1.927	1.993	Fluoride
2	6.743	0.2932	1.432	3.003	Chloride
3	8.942	0.6627	2.514	2.996	Nitrite
4	11.065	0.4373	1.402	10.047	Bromide
5	12.815	0.6071	1.700	2.489	Nitrate
6	14.778	0.4325	0.925	5.145	Phosphate
7	16.965	1.1056	2.302	14.749	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2024-10-04 13:04:49 UTC-4
Method IC1-091624
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.27 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions

