


LB132773

SFAM01.1

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.3020e-004 slope

C= 3.8182e-002 $y=10.16x$

Rho= 0.9997904

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/b
0.00	0.000	0.037	0.037	-5	0.000	-5					
0.05	0.050					442					
0.20	0.200	0.229	0.029	828	0.0 %	828	0				
2.50	2.500	2.479	-0.021	10601	0.0 %	10601					
5.00	5.000	4.976	-0.024	21448	0.0 %	21448					
7.50	7.500	7.366	-0.134	31832	0.0 %	31832					
10.0	10.000	10.114	0.114	43770	0.0 %	43770					

g/b

=

15

-1

0

-2

1

LB132773

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-5	-		0 SFAM01.1	PPB	10/4/2024 13:55	Std
	0.2 S01	828	-		0.2 SFAM01.1	PPB	10/4/2024 13:57	Std
	2.5 S02	10601	-		2.5 SFAM01.1	PPB	10/4/2024 14:02	Std
	5 S03	21448	-		5 SFAM01.1	PPB	10/4/2024 14:04	Std
	7.5 S04	31832	-		7.5 SFAM01.1	PPB	10/4/2024 14:06	Std
	10 S05	43770	-		10 SFAM01.1	PPB	10/4/2024 14:09	Std
ICV025	ICV025	16232	3.7748 -		SFAM01.1	PPB	10/4/2024 14:11	SMPL
ICB025	ICB025	-253	-0.0201 -		SFAM01.1	PPB	10/4/2024 14:14	SMPL
CCV074	CCV074	21278	4.9364 -		SFAM01.1	PPB	10/4/2024 14:16	SMPL
CCB074	CCB074	-263	-0.0224 -		SFAM01.1	PPB	10/4/2024 14:18	SMPL
PB163919BL	PBW919	185	0.0808 -		SFAM01.1	PPB	10/4/2024 14:23	SMPL
P4112-01	MH0BE8	312	0.11 -		SFAM01.1	PPB	10/4/2024 14:25	SMPL
P4112-02	MH0BF0	469	0.1461 -		SFAM01.1	PPB	10/4/2024 14:27	SMPL
P4112-03	MH0BF2	358	0.1206 -		SFAM01.1	PPB	10/4/2024 14:29	SMPL
P4112-04	MH0BF4	319	0.1116 -		SFAM01.1	PPB	10/4/2024 14:32	SMPL
P4112-05	MH0BF6	237	0.0927 -		SFAM01.1	PPB	10/4/2024 14:34	SMPL
P4112-06	MH0BF8	234	0.092 -		SFAM01.1	PPB	10/4/2024 14:36	SMPL
P4112-07	MH0BG0	290	0.1049 -		SFAM01.1	PPB	10/4/2024 14:39	SMPL
P4112-08	MH0BG2	474	0.1473 -		SFAM01.1	PPB	10/4/2024 14:41	SMPL
P4112-09	MH0BG4	246	0.0948 -		SFAM01.1	PPB	10/4/2024 14:43	SMPL
P4112-10	MH0BG6	322	0.1123 -		SFAM01.1	PPB	10/4/2024 14:45	SMPL
P4112-11	MH0BG8	267	0.0996 -		SFAM01.1	PPB	10/4/2024 14:48	SMPL
P4112-12	MH0BH0	256	0.0971 -		SFAM01.1	PPB	10/4/2024 14:50	SMPL
P4112-13	MH0BH2	316	0.1109 -		SFAM01.1	PPB	10/4/2024 14:52	SMPL
P4112-14	MH0BH4	481	0.1489 -		SFAM01.1	PPB	10/4/2024 14:54	SMPL
P4112-15	MH0BH4D	389	0.1277 -		SFAM01.1	PPB	10/4/2024 14:57	SMPL
P4112-16	MH0BH4S	4766	1.1353 -		SFAM01.1	PPB	10/4/2024 14:59	SMPL
P4112-17	MH0BH6	231	0.0914 -		SFAM01.1	PPB	10/4/2024 15:01	SMPL
P4112-18	MH0BH8	256	0.0971 -		SFAM01.1	PPB	10/4/2024 15:03	SMPL
P4112-19	MH0BJ0	391	0.1282 -		SFAM01.1	PPB	10/4/2024 15:06	SMPL
CCV075	CCV075	21344	4.9516 -		SFAM01.1	PPB	10/4/2024 15:08	SMPL
CCB075	CCB075	-212	-0.0106 -		SFAM01.1	PPB	10/4/2024 15:10	SMPL
P4112-20	MH0BJ2	429	0.1369 -		SFAM01.1	PPB	10/4/2024 15:12	SMPL
P4112-21	MH0BJ4	221	0.0891 -		SFAM01.1	PPB	10/4/2024 15:15	SMPL
P4112-22	MH0BJ6	220	0.0888 -		SFAM01.1	PPB	10/4/2024 15:17	SMPL
PB163920BL	PBW920	72	0.0548 -		SFAM01.1	PPB	10/4/2024 15:19	SMPL
P4113-01	MH0BE9	1122	0.2965 -		SFAM01.1	PPB	10/4/2024 15:21	SMPL
P4113-02	MH0BF1	1646	0.4171 -		SFAM01.1	PPB	10/4/2024 15:24	SMPL
P4113-03	MH0BF3	1447	0.3713 -		SFAM01.1	PPB	10/4/2024 15:26	SMPL
P4113-04	MH0BF5	655	0.189 -		SFAM01.1	PPB	10/4/2024 15:28	SMPL
P4113-05	MH0BF7	1223	0.3197 -		SFAM01.1	PPB	10/4/2024 15:31	SMPL
P4113-06	MH0BF9	1076	0.2859 -		SFAM01.1	PPB	10/4/2024 15:33	SMPL
P4113-07	MH0BG1	1228	0.3209 -		SFAM01.1	PPB	10/4/2024 15:35	SMPL
P4113-08	MH0BG3	1047	0.2792 -		SFAM01.1	PPB	10/4/2024 15:37	SMPL
P4113-09	MH0BG5	423	0.1356 -		SFAM01.1	PPB	10/4/2024 15:40	SMPL
P4113-10	MH0BG7	628	0.1827 -		SFAM01.1	PPB	10/4/2024 15:42	SMPL

LB132773

INSTRUMENT ID : CV1

P4113-11	MHOBG9	1221	0.3193 -	SFAM01.1	PPB	10/4/2024 15:44 SMPL
P4113-12	MHOBH1	992	0.2665 -	SFAM01.1	PPB	10/4/2024 15:46 SMPL
P4113-13	MHOBH3	735	0.2074 -	SFAM01.1	PPB	10/4/2024 15:49 SMPL
P4113-14	MHOBH5	778	0.2173 -	SFAM01.1	PPB	10/4/2024 15:51 SMPL
P4113-15	MHOBH5D	895	0.2442 -	SFAM01.1	PPB	10/4/2024 15:56 SMPL
P4113-16	MHOBH5S	6014	1.4226 -	SFAM01.1	PPB	10/4/2024 15:59 SMPL
CCV076	CCV076	21948	5.0907 -	SFAM01.1	PPB	10/4/2024 16:01 SMPL
CCB076	CCB076	-114	0.0119 -	SFAM01.1	PPB	10/4/2024 16:03 SMPL
P4113-17	MHOBH7	484	0.1496 -	SFAM01.1	PPB	10/4/2024 16:06 SMPL
P4113-18	MHOBH9	384	0.1266 -	SFAM01.1	PPB	10/4/2024 16:08 SMPL
P4113-19	MHOBJ1	343	0.1171 -	SFAM01.1	PPB	10/4/2024 16:10 SMPL
P4113-20	MHOBJ3	306	0.1086 -	SFAM01.1	PPB	10/4/2024 16:12 SMPL
P4113-21	MHOBJ5	316	0.1109 -	SFAM01.1	PPB	10/4/2024 16:15 SMPL
P4113-22	MHOBJ7	314	0.1105 -	SFAM01.1	PPB	10/4/2024 16:17 SMPL
PB163921BL	PBW921	-11	0.0357 -	SFAM01.1	PPB	10/4/2024 16:19 SMPL
P4114-01	MHOBJ8	810	0.2246 -	SFAM01.1	PPB	10/4/2024 16:21 SMPL
P4114-02	MHOBK0	459	0.1438 -	SFAM01.1	PPB	10/4/2024 16:24 SMPL
P4114-03	MHOBK2	377	0.125 -	SFAM01.1	PPB	10/4/2024 16:26 SMPL
P4114-04	MHOBK4	284	0.1036 -	SFAM01.1	PPB	10/4/2024 16:28 SMPL
P4114-05	MHOBK6	2053	0.5108 -	SFAM01.1	PPB	10/4/2024 16:31 SMPL
P4114-06	MHOBK8	1358	0.3508 -	SFAM01.1	PPB	10/4/2024 16:38 SMPL
P4114-07	MHOBK8D	1343	0.3473 -	SFAM01.1	PPB	10/4/2024 16:40 SMPL
P4114-08	MHOBK8S	4713	1.1231 -	SFAM01.1	PPB	10/4/2024 16:42 SMPL
P4114-09	MHOBLO	303	0.1079 -	SFAM01.1	PPB	10/4/2024 16:45 SMPL
P4114-10	MHOBLO2	296	0.1063 -	SFAM01.1	PPB	10/4/2024 16:47 SMPL
P4114-11	MHOBLO4	439	0.1392 -	SFAM01.1	PPB	10/4/2024 16:49 SMPL
P4114-12	MHOBLO6	284	0.1036 -	SFAM01.1	PPB	10/4/2024 16:51 SMPL
PB163922BL	PBW922	0	0.0382 -	SFAM01.1	PPB	10/4/2024 16:54 SMPL
CCV077	CCV077	21330	4.9484 -	SFAM01.1	PPB	10/4/2024 16:56 SMPL
CCB077	CCB077	-129	0.0085 -	SFAM01.1	PPB	10/4/2024 16:58 SMPL
P4115-01	MHOBJ9	300	0.1072 -	SFAM01.1	PPB	10/4/2024 17:00 SMPL
P4115-02	MHOBK1	275	0.1015 -	SFAM01.1	PPB	10/4/2024 17:03 SMPL
P4115-03	MHOBK3	343	0.1171 -	SFAM01.1	PPB	10/4/2024 17:05 SMPL
P4115-04	MHOBK5	300	0.1072 -	SFAM01.1	PPB	10/4/2024 17:07 SMPL
P4115-05	MHOBK7	344	0.1174 -	SFAM01.1	PPB	10/4/2024 17:10 SMPL
P4115-06	MHOBK7D	296	0.1063 -	SFAM01.1	PPB	10/4/2024 17:12 SMPL
P4115-07	MHOBK7S	5334	1.2661 -	SFAM01.1	PPB	10/4/2024 17:14 SMPL
P4115-08	MHOBK9	886	0.2421 -	SFAM01.1	PPB	10/4/2024 17:16 SMPL
P4115-09	MHOBLO1	317	0.1112 -	SFAM01.1	PPB	10/4/2024 17:19 SMPL
P4115-10	MHOBLO3	771	0.2157 -	SFAM01.1	PPB	10/4/2024 17:21 SMPL
P4115-11	MHOBLO5	408	0.1321 -	SFAM01.1	PPB	10/4/2024 17:23 SMPL
P4115-12	MHOBLO7	431	0.1374 -	SFAM01.1	PPB	10/4/2024 17:25 SMPL
P4161-08	YEOT1	351	0.119 -	SFAM01.1	PPB	10/4/2024 17:28 SMPL
CCV078	CCV078	22052	5.1146 -	SFAM01.1	PPB	10/4/2024 17:30 SMPL
CCB078	CCB078	-143	0.0053 -	SFAM01.1	PPB	10/4/2024 17:32 SMPL