

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132780

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/04/2024	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	10/04/2024	15:01
3	CAL3	1	Water	NA	NA	20.3	10.02	10/04/2024	15:05
4	ICV	1	Water	NA	NA	20.3	7.02	10/04/2024	15:10
5	CCV1	1	Water	NA	NA	20.3	2.01	10/04/2024	15:15
6	P4321-04	1	Solid	20.02	20	23.4	8.30	10/04/2024	15:25
7	P4321-04DUP	1	Solid	20.03	20	23.5	8.31	10/04/2024	15:26
8	P4323-01	1	Solid	20.01	20	22.7	6.87	10/04/2024	15:30
9	P4324-01	1	Solid	20.02	20	23.7	5.83	10/04/2024	15:35
10	P4324-02	1	Solid	20.3	20	23.6	6.53	10/04/2024	15:44
11	P4326-01	1	Solid	20.02	20	23.6	6.80	10/04/2024	15:50
12	CCV2	1	Water	NA	NA	20.3	12.02	10/04/2024	15:55