



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/7/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:10
In Date: 10/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132785

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4293-01	MYD9Z9	1	1.18	8.50	9.68	9.57	98.7	
P4293-02	MYDA00	2	1.18	8.34	9.52	9.38	98.3	
P4293-03	MYDA01	3	1.18	8.60	9.78	9.66	98.6	
P4293-04	MYDA02	4	1.18	8.50	9.68	9.62	99.3	
P4293-05	MYDA03	5	1.19	8.65	9.84	9.75	99.0	
P4293-06	MYDA04	6	1.16	8.50	9.66	9.48	97.9	
P4293-07	MYDA05	7	1.17	8.57	9.74	9.61	98.5	
P4293-08	MYDA05D	8	1.17	8.57	9.74	9.61	98.5	
P4293-09	MYDA05S	9	1.17	8.57	9.74	9.61	98.5	
P4293-10	MYDA06	10	1.18	8.54	9.72	9.63	98.9	
P4293-11	MYDA07	11	1.19	8.63	9.82	9.68	98.4	
P4293-12	MYDA08	12	1.19	8.51	9.7	9.58	98.6	
P4293-13	MYDA09	13	1.16	8.65	9.81	9.7	98.7	
P4293-14	MYDA10	14	1.18	8.50	9.68	9.62	99.3	
P4293-15	MYDA11	15	1.18	8.55	9.73	9.63	98.8	
P4293-16	MYDA12	16	1.17	8.55	9.72	9.68	99.5	
P4293-17	MYDA13	17	1.16	8.40	9.56	9.39	98.0	
P4293-18	MYDA14	18	1.17	8.65	9.82	9.69	98.5	
P4293-19	MYDA15	19	1.15	8.76	9.91	9.86	99.4	
P4293-20	MYDA16	20	1.15	8.79	9.94	9.77	98.1	
P4293-21	MYDA17	21	1.18	8.78	9.96	9.88	99.1	
P4293-22	MYDA18	22	1.18	8.52	9.7	9.59	98.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$