



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/7/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:35
In Date: 10/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132790

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4295-01	MYD0Z2	1	1.18	8.01	9.19	8.99	97.5	
P4295-02	MYD0Z3	2	1.18	8.44	9.62	9.31	96.3	
P4295-03	MYD0Z4	3	1.18	8.64	9.82	9.11	91.8	
P4295-04	MYD0Z5	4	1.15	8.80	9.95	8.97	88.9	
P4295-05	MYD0Z6	5	1.16	8.50	9.66	8.52	86.6	
P4295-06	MYD0Z7	6	1.18	8.42	9.6	8.48	86.7	
P4295-07	MYD0Z7D	7	1.18	8.42	9.6	8.48	86.7	
P4295-08	MYD0Z7S	8	1.18	8.42	9.6	8.48	86.7	
P4295-09	MYD0Z8	9	1.14	8.79	9.93	9.13	90.9	
P4295-10	MYD0Z9	10	1.16	8.44	9.6	8.49	86.8	
P4295-11	MYD100	11	1.14	8.84	9.98	9.77	97.6	
P4295-12	MYD101	12	1.18	8.50	9.68	9.47	97.5	
P4295-13	MYD102	13	1.15	8.71	9.86	9.31	93.7	
P4295-14	MYD103	14	1.15	8.42	9.57	8.76	90.4	
P4295-15	MYD104	15	1.18	8.59	9.77	8.9	89.9	
P4295-16	MYD105	16	1.19	8.56	9.75	8.57	86.2	
P4295-17	MYDA38	17	1.15	8.77	9.92	9.87	99.4	
P4295-18	MYDA39	18	1.14	8.80	9.94	9.83	98.8	
P4295-19	MYDA40	19	1.15	8.51	9.66	9.55	98.7	
P4295-20	MYDA41	20	1.18	8.42	9.6	9.53	99.2	
P4295-21	MYDBB3	21	1.18	8.52	9.7	9.44	96.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$