



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132798

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4299-01	MYDA93	1	1.18	8.51	9.69	9.54	98.2	
P4299-02	MYDA94	2	1.15	8.77	9.92	9.72	97.7	
P4299-03	MYDA95	3	1.15	8.57	9.72	9.54	97.9	
P4299-04	MYDA96	4	1.12	8.76	9.88	9.74	98.4	
P4299-05	MYDA97	5	1.18	8.34	9.52	9.38	98.3	
P4299-06	MYDA98	6	1.15	8.71	9.86	9.7	98.2	
P4299-07	MYDA99	7	1.18	8.40	9.58	9.56	99.8	
P4299-08	MYDAA0	8	1.15	8.73	9.88	9.75	98.5	
P4299-09	MYDAA1	9	1.14	8.70	9.84	9.62	97.5	
P4299-10	MYDAA2	10	1.15	8.40	9.55	9.43	98.6	
P4299-11	MYDAA3	11	1.18	8.43	9.61	9.47	98.3	
P4299-12	MYDAA4	12	1.15	8.82	9.97	9.83	98.4	
P4299-13	MYDAA5	13	1.18	8.57	9.75	9.68	99.2	
P4299-14	MYDAA6	14	1.18	8.56	9.74	9.57	98.0	
P4299-15	MYDAA7	15	1.15	8.45	9.6	9.41	97.8	
P4299-16	MYDAA8	16	1.15	8.43	9.58	9.44	98.3	
P4299-17	MYDAA9	17	1.14	8.63	9.77	9.66	98.7	
P4299-18	MYDAB0	18	1.13	8.60	9.73	9.57	98.1	
P4299-19	MYDAB1	19	1.12	8.38	9.5	9.47	99.6	
P4299-20	MYDAB1D	20	1.12	8.38	9.5	9.47	99.6	
P4299-21	MYDAB1S	21	1.12	8.38	9.5	9.47	99.6	
P4299-22	MYDAB2	22	1.13	8.44	9.57	9.52	99.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$