



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:20
In Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132799

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4301-01	MYDAB3	1	1.14	8.40	9.54	9.47	99.2	
P4301-02	MYDAB4	2	1.13	8.83	9.96	9.87	99.0	
P4301-03	MYDAB5	3	1.11	8.47	9.58	9.47	98.7	
P4301-04	MYDAB6	4	1.15	8.83	9.98	9.87	98.8	
P4301-05	MYDAB7	5	1.13	8.67	9.8	9.72	99.1	
P4301-06	MYDAB8	6	1.14	8.47	9.61	9.55	99.3	
P4301-07	MYDAB9	7	1.12	8.63	9.75	9.67	99.1	
P4301-08	MYDAC0	8	1.13	8.39	9.52	9.44	99.0	
P4301-09	MYDAC1	9	1.13	8.71	9.84	9.77	99.2	
P4301-10	MYDAC1D	10	1.13	8.71	9.84	9.77	99.2	
P4301-11	MYDAC1S	11	1.13	8.71	9.84	9.77	99.2	
P4301-12	MYDAC2	12	1.14	8.52	9.66	9.59	99.2	
P4301-13	MYDAC3	13	1.14	8.43	9.57	9.52	99.4	
P4301-14	MYDAC4	14	1.15	8.48	9.63	9.56	99.2	
P4301-15	MYDAC5	15	1.13	8.54	9.67	9.63	99.5	
P4301-16	MYDAC6	16	1.13	8.57	9.7	9.51	97.8	
P4301-17	MYDAC7	17	1.14	8.72	9.86	9.71	98.3	
P4301-18	MYDAC8	18	1.14	8.39	9.53	9.42	98.7	
P4301-19	MYDAC9	19	1.12	8.61	9.73	9.7	99.7	
P4301-20	MYDAD0	20	1.14	8.68	9.82	9.77	99.4	
P4301-21	MYDAD1	21	1.14	8.82	9.96	9.87	99.0	
P4301-22	MYDAD2	22	1.12	8.72	9.84	9.8	99.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$