



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132800

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4302-01	MYDAD3	1	1.13	8.65	9.78	9.72	99.3	
P4302-02	MYDAD4	2	1.12	8.48	9.6	9.5	98.8	
P4302-03	MYDAD5	3	1.13	8.40	9.53	9.34	97.7	
P4302-04	MYDAD6	4	1.13	8.39	9.52	9.44	99.0	
P4302-05	MYDAD7	5	1.13	8.62	9.75	9.6	98.3	
P4302-06	MYDAD8	6	1.13	8.66	9.79	9.65	98.4	
P4302-07	MYDAD9	7	1.12	8.55	9.67	9.63	99.5	
P4302-08	MYDAE0	8	1.12	8.51	9.63	9.56	99.2	
P4302-09	MYDAE1	9	1.14	8.52	9.66	9.57	98.9	
P4302-10	MYDAE1D	10	1.14	8.52	9.66	9.57	98.9	
P4302-11	MYDAE1S	11	1.14	8.52	9.66	9.57	98.9	
P4302-12	MYDAE2	12	1.13	8.56	9.69	9.51	97.9	
P4302-13	MYDAE3	13	1.15	8.81	9.96	9.87	99.0	
P4302-14	MYDAE4	14	1.17	8.59	9.76	9.6	98.1	
P4302-15	MYDAE5	15	1.16	8.56	9.72	9.69	99.6	
P4302-16	MYDAE6	16	1.14	8.49	9.63	9.55	99.1	
P4302-17	MYDAE8	17	1.13	8.38	9.51	9.11	95.2	
P4302-18	MYDAE9	18	1.13	8.53	9.66	9.61	99.4	
P4302-19	MYDAF0	19	1.13	8.83	9.96	9.8	98.2	
P4302-20	MYDAF1	20	1.14	8.42	9.56	9.41	98.2	
P4302-21	MYDAF2	21	1.14	8.58	9.72	9.63	99.0	
P4302-22	MYDAF3	22	1.13	8.41	9.54	9.51	99.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$