



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:20
In Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132804

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4305-01	MYDAJ5	1	1.14	8.84	9.98	9.9	99.1	
P4305-02	MYDAJ5D	2	1.14	8.84	9.98	9.9	99.1	
P4305-03	MYDAJ5S	3	1.14	8.84	9.98	9.9	99.1	
P4305-04	MYDAK0	4	1.15	8.82	9.97	9.82	98.3	
P4305-05	MYDAK1	5	1.15	8.54	9.69	9.55	98.4	
P4305-06	MYDAK2	6	1.16	8.51	9.67	9.59	99.1	
P4305-07	MYDAK3	7	1.15	8.40	9.55	9.48	99.2	
P4305-08	MYDAK4	8	1.18	8.53	9.71	9.65	99.3	
P4305-09	MYDAK5	9	1.13	8.80	9.93	9.85	99.1	
P4305-10	MYDAK6	10	1.18	8.36	9.54	9.47	99.2	
P4305-11	MYDAK7	11	1.18	8.65	9.83	9.72	98.7	
P4305-12	MYDAK8	12	1.18	8.43	9.61	9.55	99.3	
P4305-13	MYDAK9	13	1.15	8.34	9.49	9.44	99.4	
P4305-14	MYDAL0	14	1.18	8.45	9.63	9.56	99.2	
P4305-15	MYDAL1	15	1.16	8.50	9.66	9.57	98.9	
P4305-16	MYDAL2	16	1.17	8.38	9.55	9.42	98.4	
P4305-17	MYDAL3	17	1.16	8.51	9.67	9.56	98.7	
P4305-18	MYDAL4	18	1.15	8.78	9.93	9.87	99.3	
P4305-19	MYDAL5	19	1.17	8.38	9.55	9.44	98.7	
P4305-20	MYDAL6	20	1.16	8.54	9.7	9.61	98.9	
P4305-21	MYDAL7	21	1.17	8.49	9.66	9.54	98.6	
P4305-22	MYDAL8	22	1.15	8.48	9.63	9.51	98.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$