

LB132806


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4972E



A= 0.0000e+000

B= 1.9762e-004

C= 1.0189e-002

Rho= 0.9994339

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	93	0.000	93					-
0.2	0.200	0.198	-0.002	950	0.0 %	950					-1
2.5	2.500	2.380	-0.120	11990	0.0 %	11990					-5
5.0	5.000	5.035	0.035	25426	0.0 %	25426					-1
7.5	7.500	7.722	0.222	39024	0.0 %	39024					3
10.0	10.000	9.837	-0.163	49725	0.0 %	49725					-2

LB132806 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Oct 2024 13:31:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7471B	PPB	07 Oct 2024 13:42:11	S	Std
0.2 - 1	50.2	950	-	0.2000	7471B	PPB	07 Oct 2024 13:44:28	S	Std
2.5 - 1	52.5	11990	-	2.5000	7471B	PPB	07 Oct 2024 13:46:45	S	Std
5.0 - 1	55	25426	-	5.0000	7471B	PPB	07 Oct 2024 13:49:01	S	Std
7.5 - 1	57.5	39024	-	7.5000	7471B	PPB	07 Oct 2024 13:51:20	S	Std
10.0 - 1	50	49725	-	10.0000	7471B	PPB	07 Oct 2024 13:56:21	S	Std
ICV26 - 1	ICV26	19483	3.8604	-7471B	PPB	PPB	07 Oct 2024 13:59:10	U	SMPL
ICB26 - 1	ICB26	-242	-0.0376	-7471B	PPB	PPB	07 Oct 2024 14:01:26	U	SMPL
CCV79 - 1	CCV79	26021	5.1525	-7471B	PPB	PPB	07 Oct 2024 14:03:44	U	SMPL
CCB79 - 1	CCB79	-275	-0.0442	-7471B	PPB	PPB	07 Oct 2024 14:05:59	U	SMPL
CRA - 1	CRA	902	0.1884	-7471B	PPB	PPB	07 Oct 2024 14:08:16	U	SMPL
HighStd - 1	HighStd	49405	9.7736	-7471B	PPB	PPB	07 Oct 2024 14:10:31	U	SMPL
ChkStd - 1	ChkStd	33126	6.5566	-7471B	PPB	PPB	07 Oct 2024 14:12:46	U	SMPL
PB163951BL - 1	PBS	-522	-0.0930	-7471B	PPB	PPB	07 Oct 2024 14:15:09	U	SMPL
PB163951BS - 1	LCSS	17393	3.4474	-7471B	PPB	PPB	07 Oct 2024 14:17:31	U	SMPL
P4270-01 - 1	IB-2	822339	HIGH	-7471B	PPB	PPB	07 Oct 2024 14:19:47	X	
P4273-01 - 1	AU-05-100224	4158	0.8319	-7471B	PPB	PPB	07 Oct 2024 14:21:37	U	SMPL
P4273-01DUP - 1	AU-05-100224DUP	4793	0.9574	-7471B	PPB	PPB	07 Oct 2024 14:24:35	U	SMPL
P4273-01MS - 1	AU-05-100224MS	25185	4.9873	-7471B	PPB	PPB	07 Oct 2024 14:26:54	U	SMPL
P4273-01MSD - 1	AU-05-100224MSD	23932	4.7396	-7471B	PPB	PPB	07 Oct 2024 14:29:12	U	SMPL
CCV80 - 1	CCV80	25898	5.1282	-7471B	PPB	PPB	07 Oct 2024 14:31:35	U	SMPL
CCB80 - 1	CCB80	-368	-0.0625	-7471B	PPB	PPB	07 Oct 2024 14:33:54	U	SMPL
P4276-01 - 1	VNJ-208	2653	0.5345	-7471B	PPB	PPB	07 Oct 2024 14:36:11	U	SMPL
P4277-01 - 1	BU-03-10022024	3347	0.6716	-7471B	PPB	PPB	07 Oct 2024 14:38:26	U	SMPL
P4278-01 - 1	BUR-1300	281	0.0657	-7471B	PPB	PPB	07 Oct 2024 14:40:44	U	SMPL
P4288-01 - 1	HR-02-100324	17832	3.5342	-7471B	PPB	PPB	07 Oct 2024 14:43:01	U	SMPL
P4288-03 - 1	HR-03-100324	31448	6.2249	-7471B	PPB	PPB	07 Oct 2024 14:45:17	U	SMPL
P4290-01 - 1	IB-4	25	0.0151	-7471B	PPB	PPB	07 Oct 2024 14:47:36	U	SMPL
P4321-01 - 1	IB-6	97169	19.2127	-7471B	PPB	PPB	07 Oct 2024 14:49:56	U	SMPL
P4324-01 - 1	LAW-24-00169-00170	3984	0.7975	-7471B	PPB	PPB	07 Oct 2024 14:52:13	U	SMPL
P4324-02 - 1	LAW-24-00166	-46	0.0011	-7471B	PPB	PPB	07 Oct 2024 14:54:44	U	SMPL
CCV81 - 1	CCV81	26134	5.1748	-7471B	PPB	PPB	07 Oct 2024 14:57:02	U	SMPL
CCB81 - 1	CCB81	-221	-0.0335	-7471B	PPB	PPB	07 Oct 2024 14:59:17	U	SMPL
P4270-01DLX10 - 1		164923	32.6023	-7471B	PPB	PPB	07 Oct 2024 15:01:36	U	SMPL
P4321-01DLX5 - 1		20626	4.0863	-7471B	PPB	PPB	07 Oct 2024 15:03:51	U	SMPL
P4273-01LX5 - 1		706	0.1497	-7471B	PPB	PPB	07 Oct 2024 15:06:23	U	SMPL
P4273-01A - 1		24548	4.8614	-7471B	PPB	PPB	07 Oct 2024 15:08:42	U	SMPL
P4270-01DL2X50 - 1		31887	6.3117	-7471B	PPB	PPB	07 Oct 2024 15:12:05	U	SMPL
CCV82 - 1	CCV82	26155	5.1789	-7471B	PPB	PPB	07 Oct 2024 15:14:20	U	SMPL
CCB82 - 1	CCB82	-290	-0.0471	-7471B	PPB	PPB	07 Oct 2024 15:16:39	U	SMPL