

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132813

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/07/2024	14:00
2	P4285-03		1	12	100.00	>212.0	>212.0	10/07/2024	14:30
3	P4285-03DUP		1	12	100.00	>212.0	>212.0	10/07/2024	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66132813

WorkList Name : FP-10-7

WorkList ID : 184188

Department : Wet-Chemistry

Date : 10-07-2024 09:05:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4285-03	WATER-TOTES	Water	Flash Point	Cool 4 deg C	PSEG03	J51	10/03/2024	1010B

Date/Time 10/07/2024 13:50
Raw Sample Received by: RM (wc)
Raw Sample Relinquished by: JH (wc)

Date/Time 10/07/2024 15:40
Raw Sample Received by: JH (wc)
Raw Sample Relinquished by: RM (wc)