



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/8/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/07/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132814

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4309-01	MYDAQ6	1	1.14	8.55	9.69	9.47	97.4	
P4309-02	MYDAQ6D	2	1.14	8.55	9.69	9.47	97.4	
P4309-03	MYDAQ6S	3	1.14	8.55	9.69	9.47	97.4	
P4309-04	MYDAR4	4	1.18	8.71	9.89	9.68	97.6	
P4309-05	MYDAR5	5	1.12	8.74	9.86	9.75	98.7	
P4309-06	MYDAR6	6	1.18	8.72	9.9	9.78	98.6	
P4309-07	MYDAR7	7	1.15	8.43	9.58	9.45	98.5	
P4309-08	MYDAR8	8	1.18	8.46	9.64	9.47	98.0	
P4309-09	MYDAR9	9	1.18	8.56	9.74	9.58	98.1	
P4309-10	MYDAS0	10	1.15	8.61	9.76	9.6	98.1	
P4309-11	MYDAS1	22	1.13	8.62	9.75	9.49	97.0	
P4309-12	MYDAS2	11	1.18	8.42	9.6	9.19	95.1	
P4309-13	MYDAS3	12	1.17	8.41	9.58	9.28	96.4	
P4309-14	MYDAS4	13	1.18	8.72	9.9	9.87	99.7	
P4309-15	MYDAS5	14	1.15	8.59	9.74	9.63	98.7	
P4309-16	MYDAS6	15	1.13	8.80	9.93	9.68	97.2	
P4309-17	MYDAS7	16	1.12	8.60	9.72	9.63	99.0	
P4309-18	MYDAS8	17	1.18	8.67	9.85	9.44	95.3	
P4309-19	MYDAS9	18	1.14	8.60	9.74	9.6	98.4	
P4309-20	MYDAT0	19	1.18	8.62	9.8	9.54	97.0	
P4309-21	MYDAT1	20	1.15	8.67	9.82	9.57	97.1	
P4309-22	MYDAT2	21	1.18	8.53	9.71	9.35	95.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$