

LB132820


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

50597



A= 0.0000e+000

B= 1.9750e-004 *slope*C= -1.7661e-002 *y-intercept*

Rho= 0.9999162

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/11
0.00	0.000	-0.012	-0.012	30	0.000	30					
0.05	0.050					442	0				
0.20	0.200	0.207	0.007	1138	0.0 %	1138					
2.50	2.500	2.537	0.037	12937	0.0 %	12937					
5.00	5.000	4.920	-0.080	25003	0.0 %	25003					
7.50	7.500	7.572	0.072	38428	0.0 %	38428					
10.0	10.000	9.975	-0.025	50597	0.0 %	50597					

LB132820

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	30	-		0 SFAM01.1	PPB	10/8/2024 10:08	Std
	0.2 S01	1138	-		0.2 SFAM01.1	PPB	10/8/2024 10:10	Std
	2.5 S02	12937	-		2.5 SFAM01.1	PPB	10/8/2024 10:13	Std
	5 S03	25003	-		5 SFAM01.1	PPB	10/8/2024 10:15	Std
	7.5 S04	38428	-		7.5 SFAM01.1	PPB	10/8/2024 10:17	Std
	10 S05	50597	-		10 SFAM01.1	PPB	10/8/2024 10:19	Std
ICV027	ICV027	18886	3.7123 -		SFAM01.1	PPB	10/8/2024 10:23	SMPL
ICB027	ICB027	-173	-0.0518 -		SFAM01.1	PPB	10/8/2024 10:25	SMPL
CCV082	CCV082	25184	4.9561 -		SFAM01.1	PPB	10/8/2024 10:27	SMPL
CCB082	CCB082	-180	-0.0532 -		SFAM01.1	PPB	10/8/2024 10:29	SMPL
PB163975BL	PBS975	67	-0.0044 -		SFAM01.1	PPB	10/8/2024 10:32	SMPL
P4161-01	YE0S6	2950	0.565 -		SFAM01.1	PPB	10/8/2024 10:34	SMPL
P4161-02	YE0S7	2792	0.5338 -		SFAM01.1	PPB	10/8/2024 10:36	SMPL
P4161-03	YE0S7D	2469	0.47 -		SFAM01.1	PPB	10/8/2024 10:39	SMPL
P4161-04	YE0S7S	14600	2.8658 -		SFAM01.1	PPB	10/8/2024 10:44	SMPL
P4161-05	YE0S8	1509	0.2804 -		SFAM01.1	PPB	10/8/2024 10:46	SMPL
P4161-06	YE0S9	1228	0.2249 -		SFAM01.1	PPB	10/8/2024 10:48	SMPL
P4161-07	YE0T0	1355	0.2499 -		SFAM01.1	PPB	10/8/2024 10:50	SMPL
CCV083	CCV083	25144	4.9482 -		SFAM01.1	PPB	10/8/2024 10:53	SMPL
CCB083	CCB083	-106	-0.0386 -		SFAM01.1	PPB	10/8/2024 10:55	SMPL