


LB(9282)

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.2040e-004

C= 1.2641e-002

Rho= 0.9994826

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	113	0.000	113	0				20
0.2	0.200	0.239	0.039	1025	0.0 %	1025					2
2.5	2.500	2.544	0.044	11480	0.0 %	11480					2
5.0	5.000	4.898	-0.102	22157	0.0 %	22157					-2
7.5	7.500	7.305	-0.195	33076	0.0 %	33076					-3
10.0	10.000	10.177	0.177	46101	0.0 %	46101					2

LB132821 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Oct 2024 12:11:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	113	-	0.0000	7470A	PPB	08 Oct 2024 12:18:17	S	Std
0.2 - 1	50-2	1025	-	0.2000	7470A	PPB	08 Oct 2024 12:20:33	S	Std
2.5 - 1	52.5	11480	-	2.5000	7470A	PPB	08 Oct 2024 12:22:50	S	Std
5.0 - 1	55	22157	-	5.0000	7470A	PPB	08 Oct 2024 12:25:07	S	Std
7.5 - 1	59.5	33076	-	7.5000	7470A	PPB	08 Oct 2024 12:27:23	S	Std
10.0 - 1	514	46101	-	10.0000	7470A	PPB	08 Oct 2024 12:35:27	S	Std
ICV28 - 1	ICV28	16466	3.6431	-	7470A	PPB	08 Oct 2024 12:38:11	U	SMPL
ICB28 - 1	ICB28	-246	-0.0416	-	7470A	PPB	08 Oct 2024 12:40:27	U	SMPL
CCV84 - 1	CCV84	21106	4.6661	-	7470A	PPB	08 Oct 2024 12:42:44	U	SMPL
CCB84 - 1	CCB84	-272	-0.0473	-	7470A	PPB	08 Oct 2024 12:45:00	U	SMPL
CRA - 1	CRA	916	0.2146	-	7470A	PPB	08 Oct 2024 12:47:17	U	SMPL
HighStd - 1	HighStd	45988	10.1521	-	7470A	PPB	08 Oct 2024 12:49:32	U	SMPL
ChkStd - 1	ChkStd	30652	6.7708	-	7470A	PPB	08 Oct 2024 12:51:48	U	SMPL
PB163956BL - 1	PBW	-403	-0.0762	-	7470A	PPB	08 Oct 2024 12:54:06	U	SMPL
PB163956BS - 1	LCSW	17886	3.9561	-	7470A	PPB	08 Oct 2024 12:56:24	U	SMPL
P4290-04 - 1	IB-4	-271	-0.0471	-	7470A	PPB	08 Oct 2024 12:58:40	U	SMPL
P4290-04DUP - 1	IB-4DUP	-80	-0.0050	-	7470A	PPB	08 Oct 2024 13:00:58	U	SMPL
P4290-04MS - 1	IB-4MS	17525	3.8766	-	7470A	PPB	08 Oct 2024 13:03:14	U	SMPL
P4290-04MSD - 1	IB-4MSD	18892	4.1780	-	7470A	PPB	08 Oct 2024 13:05:31	U	SMPL
P4321-04 - 1	IB-6	-49	0.0018	-	7470A	PPB	08 Oct 2024 13:07:48	U	SMPL
CCV85 - 1	CCV85	22713	5.0204	-	7470A	PPB	08 Oct 2024 13:10:05	U	SMPL
CCB85 - 1	CCB85	-206	-0.0328	-	7470A	PPB	08 Oct 2024 13:12:22	U	SMPL
P4323-01 - 1	MOO-23-00323	8	0.0144	-	7470A	PPB	08 Oct 2024 13:14:39	U	SMPL
P4324-02 - 1	LAW-24-00166	-20	0.0082	-	7470A	PPB	08 Oct 2024 13:16:54	U	SMPL
P4326-01 - 1	VNJ-214	-50	0.0016	-	7470A	PPB	08 Oct 2024 13:19:09	U	SMPL
PB163967BL - 1	PBW	-80	-0.0050	-	7470A	PPB	08 Oct 2024 13:21:24	U	SMPL
P4254-05 - 1	MH-147	1533	0.3506	-	7470A	PPB	08 Oct 2024 13:28:13	U	SMPL
P4269-02 - 1	COMP	-84	-0.0059	-	7470A	PPB	08 Oct 2024 13:30:37	U	SMPL
P4269-02DUP - 1	COMPDUP	357	0.0914	-	7470A	PPB	08 Oct 2024 13:32:56	U	SMPL
P4269-02MS - 1	COMPMS	15402	3.4085	-	7470A	PPB	08 Oct 2024 13:37:14	U	SMPL
CCV86 - 1	CCV86	22437	4.9596	-	7470A	PPB	08 Oct 2024 13:39:30	U	SMPL
CCB86 - 1	CCB86	-276	-0.0482	-	7470A	PPB	08 Oct 2024 13:41:46	U	SMPL
P4269-02MSD - 1	COMPMSD	16339	3.6151	-	7470A	PPB	08 Oct 2024 13:44:03	U	SMPL
P4274-01 - 1	222254	-186	-0.0284	-	7470A	PPB	08 Oct 2024 13:46:19	U	SMPL
P4285-03 - 1	WATER-TOTES	-2	0.0122	-	7470A	PPB	08 Oct 2024 13:48:34	U	SMPL
P4341-01 - 1	10324-0	439	0.1094	-	7470A	PPB	08 Oct 2024 13:50:50	U	SMPL
P4342-01 - 1	PORT-ELIZABETH-1201	26912	5.9462	-	7470A	PPB	08 Oct 2024 13:53:06	U	SMPL
P3847-07 - 1	PT-HG-WP	111566	24.6107	-	7470A	PPB	08 Oct 2024 13:55:22	U	SMPL
PB163956TB - 1	PB163956TB	57	0.0252	-	7470A	PPB	08 Oct 2024 14:02:53	U	SMPL
CCV87 - 1	CCV87	22678	5.0127	-	7470A	PPB	08 Oct 2024 14:05:09	U	SMPL
CCB87 - 1	CCB87	-134	-0.0169	-	7470A	PPB	08 Oct 2024 14:07:24	U	SMPL
P4290-04LX5 - 1		-20	0.0082	-	7470A	PPB	08 Oct 2024 14:09:41	U	SMPL
P4290-04A - 1		18790	4.1555	-	7470A	PPB	08 Oct 2024 14:11:56	U	SMPL
P4269-02LX5 - 1		-158	-0.0222	-	7470A	PPB	08 Oct 2024 14:14:12	U	SMPL
P4269-02A - 1		16690	3.6925	-	7470A	PPB	08 Oct 2024 14:16:28	U	SMPL
PB163967BS - 1	LCSW	18103	4.0040	-	7470A	PPB	08 Oct 2024 14:18:43	U	SMPL
P3847-07DLX5 - 1		22959	5.0746	-	7470A	PPB	08 Oct 2024 14:25:36	U	SMPL
CCV88 - 1	CCV88	23089	5.1033	-	7470A	PPB	08 Oct 2024 14:27:54	U	SMPL
CCB88 - 1	CCB88	-274	-0.0478	-	7470A	PPB	08 Oct 2024 14:30:12	U	SMPL