



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 10/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132822

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4329-01	A4E78	1	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4329-02	A4E79	2	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4329-03	A4EG4	3	1.15	8.74	9.89	4.64	39.9	
P4329-04	A4EG5	4	1.15	8.69	9.84	8.37	83.1	
P4329-05	A4EG8	5	1.15	8.61	9.76	4.99	44.6	
P4329-06	A4EN3	6	1.15	8.81	9.96	7.29	69.7	
P4329-07	A4EN4	7	1.17	8.60	9.77	5.9	55.0	
P4329-08	A4EN5	8	1.16	8.48	9.64	7.92	79.7	
P4329-09	A4ER1	9	1.18	8.79	9.97	6.21	57.2	
P4329-10	A4ER2	10	1.16	8.47	9.63	4.97	45.0	
P4329-11	A4ER3	11	1.17	8.58	9.75	7.07	68.8	
P4329-13	VHBLK001	12	1.00	1.00	2.00	2.00	100.0	vhblk
P4329-15	A4FB4	13	1.15	8.83	9.98	8.85	87.2	
P4329-16	A4FB5	14	1.15	8.83	9.98	8.19	79.7	
P4329-17	A4FB6	15	1.14	8.80	9.94	8.52	83.9	
P4329-18	A4FB7	16	1.18	8.79	9.97	8.62	84.6	
P4329-19	A4FB7MS	17	1.18	8.79	9.97	8.62	84.6	
P4329-20	A4FB7MSD	18	1.18	8.79	9.97	8.62	84.6	
P4329-21	A4FB8	19	1.18	8.59	9.77	8.49	85.1	
P4329-22	A4FC6	20	1.19	8.59	9.78	8.71	87.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$