

LB132826

245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40860

A= 0.0000e+000

B= 2.4716e-004

C= 1.2461e-002

Rho= 0.9998280

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.020	0.020	31	0.000	31					10
0.05	0.050	0.055	0.005	173	0.0 %	173	0				9
0.2	0.200	0.219	0.019	834	0.0 %	834					6
2.5	2.500	2.512	0.012	10115	0.0 %	10115					-1
5.0	5.000	4.970	-0.030	20056	0.0 %	20056					-2
7.5	7.500	7.363	-0.137	29738	0.0 %	29738					1
10.0	10.000	10.111	0.111	40860	0.0 %	40860					

LB132826 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Oct 2024 15:37:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	245.1	PPB	08 Oct 2024 15:48:22	S	Std
0.05 - 1	50.05	173	-	0.0500	245.1	PPB	08 Oct 2024 15:52:53	S	Std
0.2 - 1	50.2	834	-	0.2000	245.1	PPB	08 Oct 2024 15:55:10	S	Std
2.5 - 1	52.5	10115	-	2.5000	245.1	PPB	08 Oct 2024 15:57:26	S	Std
5.0 - 1	55	20056	-	5.0000	245.1	PPB	08 Oct 2024 15:59:42	S	Std
7.5 - 1	57.5	29738	-	7.5000	245.1	PPB	08 Oct 2024 16:01:59	S	Std
10.0 - 1	510	40860	-	10.0000	245.1	PPB	08 Oct 2024 16:04:16	S	Std
ICV29 - 1	ICV29	16034	3.9754	-245.1		PPB	08 Oct 2024 16:09:49	U	SMPL
ICB29 - 1	ICB29	-163	-0.0278	-245.1		PPB	08 Oct 2024 16:12:03	U	SMPL
CCV89 - 1	CCV89	19915	4.9347	-245.1		PPB	08 Oct 2024 16:14:20	U	SMPL
CCB89 - 1	CCB89	-196	-0.0360	-245.1		PPB	08 Oct 2024 16:16:35	U	SMPL
CRA - 1	CRA	159	0.0518	-245.1		PPB	08 Oct 2024 16:18:51	U	SMPL
HighStd - 1	HighStd	42894	10.6142	-245.1		PPB	08 Oct 2024 16:23:56	U	SMPL
ChkStd - 1	ChkStd	27013	6.6890	-245.1		PPB	08 Oct 2024 16:26:11	U	SMPL
PB163988BL - 1	PBW	-320	-0.0666	-245.1		PPB	08 Oct 2024 16:28:29	U	SMPL
PB163988BS - 1	LCSW	14934	3.7036	-245.1		PPB	08 Oct 2024 16:33:26	U	SMPL
P4262-07 - 1	14B-(1-4)-COMP	4537	1.1338	-245.1		PPB	08 Oct 2024 16:37:59	U	SMPL
P4271-07 - 1	14B-(1-4)-COMP	1488	0.3802	-245.1		PPB	08 Oct 2024 16:40:25	U	SMPL
P4289-02 - 1	EFF-Waste Water	85	0.0335	-245.1		PPB	08 Oct 2024 16:42:45	U	SMPL
P4289-02DUP - 1	EFF-Waste WaterDUP	281	0.0819	-245.1		PPB	08 Oct 2024 16:45:02	U	SMPL
CCV90 - 1	CCV90	21520	5.3314	-245.1		PPB	08 Oct 2024 16:51:53	U	SMPL
CCB90 - 1	CCB90	-151	-0.0249	-245.1		PPB	08 Oct 2024 16:54:08	U	SMPL
P4289-02MS - 1	EFF-Waste WaterMS	14844	3.6813	-245.1		PPB	08 Oct 2024 16:56:26	U	SMPL
P4289-02MSD - 1	EFF-Waste WaterMSD	15315	3.7977	-245.1		PPB	08 Oct 2024 16:58:45	U	SMPL
P3846-08 - 1	PT-HG-WP	110521	27.3290	-245.1		PPB	08 Oct 2024 17:01:03	U	SMPL
P4289-02LX5 - 1		-905	-0.2112	-245.1		PPB	08 Oct 2024 17:08:02	U	SMPL
P4289-02A - 1		17367	4.3049	-245.1		PPB	08 Oct 2024 17:15:42	U	SMPL
P3846-08DLX5 - 1		21585	5.3474	-245.1		PPB	08 Oct 2024 17:17:59	U	SMPL
CCV91 - 1	CCV91	21876	5.4194	-245.1		PPB	08 Oct 2024 17:24:50	U	SMPL
CCB91 - 1	CCB91	-298	-0.0612	-245.1		PPB	08 Oct 2024 17:27:08	U	SMPL