



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/9/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:35
In Date: 10/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 10/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132828

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4353-01	CF-400-COMP-43	11	1.15	8.53	9.68	9.09	93.1	
P4353-02	CF-400-VOC-43	12	1.13	8.80	9.93	9.5	95.1	
P4353-03	CF-400-93	13	1.16	8.48	9.64	9.29	95.9	
P4353-04	CF-400-94	14	1.15	8.35	9.5	8.8	91.6	
P4353-05	CF-400-95	15	1.16	8.68	9.84	9.00	90.3	
P4355-01	PIPE-11	16	1.00	1.00	2.00	2.00	100.0	wipe sample
P4355-02	PIPE-12	17	1.00	1.00	2.00	2.00	100.0	wipe sample
P4355-03	PIPE-13	18	1.00	1.00	2.00	2.00	100.0	wipe sample
P4356-01	CONCRETE-1	19	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4356-03	CONCRETE-2	20	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4357-01	ORA-2051	1	1.00	1.00	2.00	2.00	100.0	wipe sample
P4357-02	ORA-2052	2	1.00	1.00	2.00	2.00	100.0	wipe sample
P4357-03	ORA-2053	3	1.00	1.00	2.00	2.00	100.0	wipe sample
P4357-04	ORA-1909	4	1.00	1.00	2.00	2.00	100.0	wipe sample
P4358-01	2032	5	1.15	8.78	9.93	9.52	95.3	
P4358-02	23-A	6	1.18	8.60	9.78	9.77	99.9	
P4358-03	23-B	7	1.18	8.42	9.6	9.56	99.5	
P4358-04	24-A	8	1.16	8.40	9.56	9.48	99.0	
P4358-05	24-B	9	1.18	8.50	9.68	9.66	99.8	
P4359-01	EXCAVATION-SOIL	10	1.17	8.58	9.75	8.99	91.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$