

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132834

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	10/09/2024	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	10/09/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	10/09/2024	08:17
4	ICV	1	Water	NA	NA	20.3	7.01	10/09/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	10/09/2024	08:25
6	P4341-01	1	Water	NA	NA	22.6	8.17	10/09/2024	08:33
7	P4342-01	1	Water	NA	NA	23.1	7.74	10/09/2024	08:40
8	P4354-01	1	Water	NA	NA	22.1	9.63	10/09/2024	08:48
9	P4354-01DUP	1	Water	NA	NA	22.2	9.64	10/09/2024	08:49
10	CCV2	1	Water	NA	NA	20.3	12.02	10/09/2024	08:55

WORKLIST(Hardcopy Internal Chain)

VB 132834

WorkList Name : PH P4342 WorkList ID : 184256 Department : Wet-Chemistry Date : 10-09-2024 08:01:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4341-01	C 10324-0	Water	pH	Cool 4 deg C	PSEG03	K13	10/07/2024	9040C
P4342-01	P PORT-ELIZABETH-1201	Water	pH	Cool 4 deg C	PSEG03	K13	10/07/2024	9040C
P4354-01	A 291551	Water	pH	Cool 4 deg C	PSEG03	K21	10/08/2024	9040C

Date/Time 10/09/24 08:10
Raw Sample Received by: JB WOC
Raw Sample Relinquished by: JB WOC

Date/Time 10/09/24 11:00
Raw Sample Received by: JB WOC
Raw Sample Relinquished by: JB WOC