


LB132842

7471B

INSTRUMENT ID: CVJ

Linear



A= 0.0000e+000

B= 1.8565e-004

C= -9.4089e-002

Rho= 0.9997144

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	0.047	0.047	760	0.000	760					1
0.2	0.200	0.196	-0.004	1562	0.0 %	1562					-2
2.5	2.500	2.553	0.053	14259	0.0 %	14259					2
5.0	5.000	4.810	-0.190	26413	0.0 %	26413					-4
7.5	7.500	7.568	0.068	41272	0.0 %	41272					1
10.0	10.000	10.026	0.026	54512	0.0 %	54512					0

LB132842 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Oct 2024 11:41:59

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	30	760	-	0.0000	7471B	PPB	09 Oct 2024 11:53:18	S	Std
0.2 - 1	50.2	1562	-	0.2000	7471B	PPB	09 Oct 2024 11:57:38	S	Std
2.5 - 1	32.5	14259	-	2.5000	7471B	PPB	09 Oct 2024 11:59:54	S	Std
5.0 - 1	35	26413	-	5.0000	7471B	PPB	09 Oct 2024 12:02:11	S	Std
7.5 - 1	59.5	41272	-	7.5000	7471B	PPB	09 Oct 2024 12:04:30	S	Std
10.0 - 1	310	54512	-	10.0000	7471B	PPB	09 Oct 2024 12:06:48	S	Std
ICV30 - 1	ICV30	20129	3.6429	-	7471B	PPB	09 Oct 2024 12:09:45	U	SMPL
ICB30 - 1	ICB30	139	-0.0683	-	7471B	PPB	09 Oct 2024 12:12:01	U	SMPL
CCV92 - 1	CCV92	26548	4.8346	-	7471B	PPB	09 Oct 2024 12:14:18	U	SMPL
CCB92 - 1	CCB92	136	-0.0688	-	7471B	PPB	09 Oct 2024 12:16:33	U	SMPL
CRA - 1	CRA	1459	0.1768	-	7471B	PPB	09 Oct 2024 12:18:51	U	SMPL
HighStd - 1	HighStd	53127	9.7691	-	7471B	PPB	09 Oct 2024 12:21:07	U	SMPL
ChkStd - 1	ChkStd	38534	7.0599	-	7471B	PPB	09 Oct 2024 12:23:22	U	SMPL
PB164002BL - 1	PB164002BL	308	-0.0369	-	7471B	PPB	09 Oct 2024 12:28:49	U	SMPL
PB164002BS - 1	PB164002BS	19581	3.5412	-	7471B	PPB	09 Oct 2024 12:34:59	U	SMPL
P4340-01 - 1	1B-7	121986	22.5530	-	7471B	PPB	09 Oct 2024 12:37:20	U	SMPL
P4340-05 - 1	1B-3	1391	0.1642	-	7471B	PPB	09 Oct 2024 12:39:39	U	SMPL
P4343-01 - 1	EO-03-100724	2502	0.3704	-	7471B	PPB	09 Oct 2024 12:42:19	U	SMPL
P4344-01 - 1	TR-05-100724	1178	0.1246	-	7471B	PPB	09 Oct 2024 12:44:37	U	SMPL
P4353-01 - 1	CF-400-COMP-43	2369	0.3457	-	7471B	PPB	09 Oct 2024 12:46:53	U	SMPL
CCV93 - 1	CCV93	26362	4.8001	-	7471B	PPB	09 Oct 2024 12:49:08	U	SMPL
CCB93 - 1	CCB93	57	-0.0835	-	7471B	PPB	09 Oct 2024 12:51:25	U	SMPL
P4359-01 - 1	EXCAVATION-SOIL	2647	0.3973	-	7471B	PPB	09 Oct 2024 12:53:43	U	SMPL
P4359-01DUP - 1	EXCAVATION-SOILDUP	2307	0.3342	-	7471B	PPB	09 Oct 2024 12:55:59	U	SMPL
P4359-01MS - 1	EXCAVATION-SOILMS	22975	4.1713	-	7471B	PPB	09 Oct 2024 13:04:39	U	SMPL
P4359-01MSD - 1	EXCAVATION-SOILMSD	22482	4.0798	-	7471B	PPB	09 Oct 2024 13:06:55	U	SMPL
P4359-01LX5 - 1		405	-0.0189	-	7471B	PPB	09 Oct 2024 13:09:13	U	SMPL
P4359-01A - 1		23752	4.3155	-	7471B	PPB	09 Oct 2024 13:11:31	U	SMPL
P4340-01DLX5 - 1		26745	4.8712	-	7471B	PPB	09 Oct 2024 13:13:47	U	SMPL
CCV94 - 1	CCV94	25933	4.7204	-	7471B	PPB	09 Oct 2024 13:19:05	U	SMPL
CCB94 - 1	CCB94	488	-0.0035	-	7471B	PPB	09 Oct 2024 13:21:25	U	SMPL