



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/10/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 10/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132844

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4318-01	MYDB12	1	1.15	8.65	9.8	9.53	96.9	
P4318-02	MYDB13	2	1.17	8.40	9.57	9.44	98.5	
P4318-03	MYDB14	3	1.15	8.74	9.89	9.69	97.7	
P4318-04	MYDB15	4	1.15	8.50	9.65	9.5	98.2	
P4318-05	MYDB16	5	1.17	8.40	9.57	9.35	97.4	
P4318-06	MYDB18	6	1.18	8.46	9.64	9.44	97.6	
P4318-07	MYDB19	7	1.16	8.50	9.66	9.47	97.8	
P4318-08	MYDB20	8	1.18	8.54	9.72	9.62	98.8	
P4318-09	MYDB21	9	1.15	8.48	9.63	9.39	97.2	
P4318-10	MYDB22	10	1.15	8.57	9.72	9.41	96.4	
P4318-11	MYDB63	11	1.15	8.37	9.52	9.18	95.9	
P4318-12	MYDB64	12	1.12	8.76	9.88	9.7	97.9	
P4318-13	MYDB65	13	1.15	8.40	9.55	9.26	96.5	
P4318-14	MYDB66	14	1.15	8.65	9.8	9.52	96.8	
P4318-15	MYDB67	15	1.17	8.65	9.82	9.61	97.6	
P4318-16	MYDB68	16	1.14	8.40	9.54	9.32	97.4	
P4318-17	MYDB69	17	1.14	8.40	9.54	9.3	97.1	
P4318-18	MYDB70	18	1.17	8.50	9.67	9.38	96.6	
P4318-19	MYDB71	19	1.16	8.50	9.66	9.44	97.4	
P4318-20	MYDB72	20	1.17	8.54	9.71	9.41	96.5	
P4318-21	MYDB72D	21	1.17	8.54	9.71	9.41	96.5	
P4318-22	MYDB72S	22	1.17	8.54	9.71	9.41	96.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$