

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB132846

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/09/2024	09:30
2	P4341-01		1	12	100.00	>212.0	>212.0	10/09/2024	10:00
3	P4341-01DUP		1	12	100.00	>212.0	>212.0	10/09/2024	10:30
4	P4342-01		1	11	100.00	>212.0	>212.0	10/09/2024	11:00
5	P4354-01		1	11	100.00	>212.0	>212.0	10/09/2024	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132846

WorkList Name : fp10-9

WorkList ID : 184270

Department : Wet-Chemistry

Date : 10-09-2024 08:49:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4341-01	10324-0	Water	Flash Point	Cool 4 deg C	PSEG03	K13	10/07/2024	1010B
P4342-01	PORT-ELIZABETH-1201	Water	Flash Point	Cool 4 deg C	PSEG03	K13	10/07/2024	1010B
P4354-01	291551	Water	Flash Point	Cool 4 deg C	PSEG03	K21	10/08/2024	1010B

Date/Time 10/09/24 08:50
Raw Sample Received by: 12607
Raw Sample Relinquished by: 12607

Date/Time 10/09/24 12:10
Raw Sample Received by: 12607
Raw Sample Relinquished by: 12607