


LB132850

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.1188e-004 *slope*

C= -8.9807e-003 *+ intercept*

Rho= 0.9992276

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	34	0.000	34					
0.05	0.050					442	0				
0.20	0.200	0.240	0.040	1174	0.0 %	1174					
2.50	2.500	2.501	0.001	11846	0.0 %	11846					
5.00	5.000	5.098	0.098	24105	0.0 %	24105					
7.50	7.500	7.202	-0.298	34031	0.0 %	34031					
10.0	10.000	10.161	0.161	47997	0.0 %	47997					

0.0
=
20
0
2
-4
2

LB132850
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Con	Method	Units	Date	Type
	0 S0	34	-		0 SFAM01.1	PPB	10/9/2024 15:39	Std
	0.2 S01	1174	-		0.2 SFAM01.1	PPB	10/9/2024 15:41	Std
	2.5 S02	11846	-		2.5 SFAM01.1	PPB	10/9/2024 15:44	Std
	5 S03	24105	-		5 SFAM01.1	PPB	10/9/2024 15:46	Std
	7.5 S04	34031	-		7.5 SFAM01.1	PPB	10/9/2024 15:48	Std
	10 S05	47997	-		10 SFAM01.1	PPB	10/9/2024 15:53	Std
ICV031	ICV031	17485	3.6958	-	SFAM01.1	PPB	10/9/2024 15:56	SMPL
ICB031	ICB031	-134	-0.0374	-	SFAM01.1	PPB	10/9/2024 15:58	SMPL
CCV095	CCV095	23818	5.0377	-	SFAM01.1	PPB	10/9/2024 16:01	SMPL
CCB095	CCB095	-179	-0.0469	-	SFAM01.1	PPB	10/9/2024 16:03	SMPL
PB164008BL	PBW008	129	0.0184	-	SFAM01.1	PPB	10/9/2024 16:05	SMPL
P4227-01	PMW-1(20240925)	6271	1.3197	-	SFAM01.1	PPB	10/9/2024 16:07	SMPL
P4227-02	PMW-2(20240925)	1193	0.2438	-	SFAM01.1	PPB	10/9/2024 16:10	SMPL
P4227-03	PMW-9S(20240926)	201	0.0336	-	SFAM01.1	PPB	10/9/2024 16:12	SMPL
P4227-04	PMW-9D(20240926)	207	0.0349	-	SFAM01.1	PPB	10/9/2024 16:14	SMPL
P4227-05	PMW-8S(20240926)	136	0.0198	-	SFAM01.1	PPB	10/9/2024 16:16	SMPL
P4227-06	PMW-8D(20240926)	142	0.0211	-	SFAM01.1	PPB	10/9/2024 16:19	SMPL
P4227-07	DUP(20240926)	157	0.0243	-	SFAM01.1	PPB	10/9/2024 16:21	SMPL
P4227-08	MW-2(MCUA)(20240926)	-13	-0.0117	-	SFAM01.1	PPB	10/9/2024 16:23	SMPL
P4227-09	PMW-5(20240926)	235	0.0408	-	SFAM01.1	PPB	10/9/2024 16:26	SMPL
P4227-10	MW-1(MCUA)(20240926)	20	-0.0047	-	SFAM01.1	PPB	10/9/2024 16:28	SMPL
P4227-11	PMW-3(20240927)	231	0.04	-	SFAM01.1	PPB	10/9/2024 16:30	SMPL
P4227-12	PMW-4(20240927)	147	0.0222	-	SFAM01.1	PPB	10/9/2024 16:32	SMPL
P4227-13	PMW-6(20240927)	178	0.0287	-	SFAM01.1	PPB	10/9/2024 16:35	SMPL
P4227-14	PMW-7S(20240927)	-4	-0.0098	-	SFAM01.1	PPB	10/9/2024 16:37	SMPL
P4227-15	PMW-7D(20240927)	22	-0.0043	-	SFAM01.1	PPB	10/9/2024 16:39	SMPL
P4227-16	P4227-15D	60	0.0037	-	SFAM01.1	PPB	10/9/2024 16:44	SMPL
P4227-17	P4227-15S	3772	0.7902	-	SFAM01.1	PPB	10/9/2024 16:48	SMPL
P4227-18	PMW-10(20240927)	869	0.1751	-	SFAM01.1	PPB	10/9/2024 16:51	SMPL
P4227-19	FB(20240925)	555	0.1086	-	SFAM01.1	PPB	10/9/2024 16:53	SMPL
CCV096	CCV096	24037	5.0841	-	SFAM01.1	PPB	10/9/2024 16:55	SMPL
CCB096	CCB096	-156	-0.042	-	SFAM01.1	PPB	10/9/2024 16:57	SMPL
P4227-20	FB(20240926)	30	-0.0026	-	SFAM01.1	PPB	10/9/2024 17:00	SMPL
P4227-21	FB(20240927)	44	0.0003	-	SFAM01.1	PPB	10/9/2024 17:02	SMPL
P4227-22	EB(20240927)	92	0.0105	-	SFAM01.1	PPB	10/9/2024 17:04	SMPL
CCV097	CCV097	23770	5.0275	-	SFAM01.1	PPB	10/9/2024 17:07	SMPL
CCB097	CCB097	-99	-0.03	-	SFAM01.1	PPB	10/9/2024 17:09	SMPL