

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132851

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/09/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/09/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/09/2024	16:30
4	ICV	1	Water	NA	NA	20.3	7.01	10/09/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	10/09/2024	16:37
6	P4372-01	1	Water	NA	NA	20.3	6.24	10/09/2024	16:44
7	P4372-02	1	Water	NA	NA	20.6	6.63	10/09/2024	16:50
8	P4372-02DUP	1	Water	NA	NA	20.7	6.64	10/09/2024	16:51
9	CCV2	1	Water	NA	NA	20.2	12.02	10/09/2024	17:00

WORKLIST(Hardcopy Internal Chain)

UP 132851

WorkList Name : ph p4372 WorkList ID : 184289 Department : Wet-Chemistry Date : 10-09-2024 16:09:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4372-01	412	Water	pH	Cool 4 deg C	PSEG03	K21	10/09/2024	9040C
P4372-02	415	Water	pH	Cool 4 deg C	PSEG03	K21	10/09/2024	9040C

Date/Time 10/09/24 16:15
 Raw Sample Received by: SS WOC
 Raw Sample Relinquished by: SS WOC

Date/Time 10/09/24 17:30
 Raw Sample Received by: SS WOC
 Raw Sample Relinquished by: SS WOC