


LB132856

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.6970e-004

C= -2.6340e-002

Rho= 0.9994599

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.0	0.000	-0.011	-0.011	57	0.000	57	0				1
0.2	0.200	0.217	0.017	904	0.0 %	904					8
2.5	2.500	2.458	-0.042	9213	0.0 %	9213					-2
5.0	5.000	4.946	-0.054	18436	0.0 %	18436					-1
7.5	7.500	7.744	0.244	28809	0.0 %	28809					3
10.0	10.000	9.846	-0.154	36603	0.0 %	36603					2

LB132856 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Oct 2024 10:45:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	10 Oct 2024 11:07:06	S	Std
0.2 - 1	50.2	904	-	0.2000	7470A	PPB	10 Oct 2024 11:09:23	S	Std
2.5 - 1	52.5	9213	-	2.5000	7470A	PPB	10 Oct 2024 11:11:40	S	Std
5.0 - 1	55	18436	-	5.0000	7470A	PPB	10 Oct 2024 11:13:57	S	Std
7.5 - 1	57.5	28809	-	7.5000	7470A	PPB	10 Oct 2024 11:21:25	S	Std
10.0 - 1	510	36603	-	10.0000	7470A	PPB	10 Oct 2024 11:29:44	S	Std
ICV33 - 1	ICV33	13848	3.7085	-	7470A	PPB	10 Oct 2024 11:32:39	U	SMPL
ICB33 - 1	ICB33	-211	-0.0832	-	7470A	PPB	10 Oct 2024 11:34:55	U	SMPL
CCV01 - 1	CCV01	21438	5.7556	-	7470A	PPB	10 Oct 2024 11:37:12	U	SMPL
CCB01 - 1	CCB01	0	-0.0263	-	7470A	PPB	10 Oct 2024 11:39:27	U	SMPL
CCV01 - 1	CCV01	18476	4.9567	-	7470A	PPB	10 Oct 2024 11:40:04	U	SMPL
CCB01 - 1	CCB01	-126	-0.0603	-	7470A	PPB	10 Oct 2024 11:42:19	U	SMPL
CRA - 1	CRA	960	0.2326	-	7470A	PPB	10 Oct 2024 11:44:36	U	SMPL
HighStd - 1	HighStd	36175	9.7302	-	7470A	PPB	10 Oct 2024 11:46:51	U	SMPL
ChkStd - 1	ChkStd	25388	6.8209	-	7470A	PPB	10 Oct 2024 11:49:06	U	SMPL
PB164014BL - 1	PBW	56	-0.0112	-	7470A	PPB	10 Oct 2024 11:53:57	U	SMPL
PB164014BS - 1	LCSW	17251	4.6263	-	7470A	PPB	10 Oct 2024 11:56:13	U	SMPL
P4340-04 - 1	1B-7	-194	-0.0787	-	7470A	PPB	10 Oct 2024 11:58:29	U	SMPL
P4340-08 - 1	1B-3	-58	-0.0420	-	7470A	PPB	10 Oct 2024 12:00:47	U	SMPL
P4353-06 - 1	CF-400-COMP-43	25	-0.0196	-	7470A	PPB	10 Oct 2024 12:03:04	U	SMPL
P4358-01 - 1	2032	-12	-0.0296	-	7470A	PPB	10 Oct 2024 12:05:21	U	SMPL
P4358-02 - 1	23-A	-9	-0.0288	-	7470A	PPB	10 Oct 2024 12:07:36	U	SMPL
CCV02 - 1	CCV02	19255	5.1668	-	7470A	PPB	10 Oct 2024 12:09:51	U	SMPL
CCB02 - 1	CCB02	-177	-0.0741	-	7470A	PPB	10 Oct 2024 12:12:07	U	SMPL
P4358-03 - 1	23-B	-80	-0.0479	-	7470A	PPB	10 Oct 2024 12:14:24	U	SMPL
P4358-04 - 1	24-A	318	0.0594	-	7470A	PPB	10 Oct 2024 12:16:39	U	SMPL
P4358-05 - 1	24-B	-84	-0.0490	-	7470A	PPB	10 Oct 2024 12:18:55	U	SMPL
P4359-02 - 1	EXCAVATION-SOIL	-76	-0.0468	-	7470A	PPB	10 Oct 2024 12:21:12	U	SMPL
P4359-02DUP - 1	EXCAVATION-SOILDUP	17	-0.0218	-	7470A	PPB	10 Oct 2024 12:23:28	U	SMPL
P4359-02MS - 1	EXCAVATION-SOILMS	15699	4.2077	-	7470A	PPB	10 Oct 2024 12:28:30	U	SMPL
P4359-02MSD - 1	EXCAVATION-SOILMSD	15373	4.1198	-	7470A	PPB	10 Oct 2024 12:30:46	U	SMPL
PB164015BL - 1	PBW	-82	-0.0485	-	7470A	PPB	10 Oct 2024 12:33:05	U	SMPL
PB164015BS - 1	LCSW	15742	4.2193	-	7470A	PPB	10 Oct 2024 12:35:23	U	SMPL
P4341-02 - 1	10324-0	726	0.1695	-	7470A	PPB	10 Oct 2024 12:37:38	U	SMPL
CCV03 - 1	CCV03	18136	4.8650	-	7470A	PPB	10 Oct 2024 12:42:33	U	SMPL
CCB03 - 1	CCB03	-159	-0.0692	-	7470A	PPB	10 Oct 2024 12:44:49	U	SMPL
P4341-02DUP - 1	10324-0DUP	719	0.1676	-	7470A	PPB	10 Oct 2024 12:47:06	U	SMPL
P4341-02MS - 1	10324-0MS	17667	4.7385	-	7470A	PPB	10 Oct 2024 12:52:00	U	SMPL
P4341-02MSD - 1	10324-0MSD	15173	4.0659	-	7470A	PPB	10 Oct 2024 12:54:16	U	SMPL
P4372-01 - 1	412	-215	-0.0843	-	7470A	PPB	10 Oct 2024 12:56:34	U	SMPL
P4372-02 - 1	415	-34	-0.0355	-	7470A	PPB	10 Oct 2024 12:58:50	U	SMPL
PB163981TB - 1	PB163981TB	-13	-0.0298	-	7470A	PPB	10 Oct 2024 13:01:06	U	SMPL
PB163983TB - 1	PB163983TB	-3	-0.0271	-	7470A	PPB	10 Oct 2024 13:03:23	U	SMPL
PB164010TB - 1	PB164010TB	22	-0.0204	-	7470A	PPB	10 Oct 2024 13:05:40	U	SMPL
CCV04 - 1	CCV04	19505	5.2342	-	7470A	PPB	10 Oct 2024 13:07:57	U	SMPL
CCB04 - 1	CCB04	-131	-0.0617	-	7470A	PPB	10 Oct 2024 13:10:12	U	SMPL
P4359-02LX5 - 1		-64	-0.0436	-	7470A	PPB	10 Oct 2024 13:12:29	U	SMPL
P4359-02A - 1		16793	4.5028	-	7470A	PPB	10 Oct 2024 13:14:44	U	SMPL
P4341-02LX5 - 1		92	-0.0015	-	7470A	PPB	10 Oct 2024 13:16:59	U	SMPL
P4341-02A - 1		15987	4.2854	-	7470A	PPB	10 Oct 2024 13:19:16	U	SMPL
CCV05 - 1	CCV05	18087	4.8518	-	7470A	PPB	10 Oct 2024 13:21:32	U	SMPL
CCV05 - 1	CCV05	-162	-0.0700	-	7470A	PPB	10 Oct 2024 13:23:49	U	SMPL