

LB132869


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

44917



A= 0.0000e+000

B= 2.2620e-004 slope

C= -2.1606e-002 y-intercept

Rho= 0.9993627

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.013	-0.013	40	0.000	40					
0.05	0.050					442	0				
0.20	0.200	0.237	0.037	1145	0.0 %	1145					
2.50	2.500	2.503	0.003	11159	0.0 %	11159					
5.00	5.000	5.104	0.104	22658	0.0 %	22658					
7.50	7.500	7.231	-0.269	32061	0.0 %	32061					
10.0	10.000	10.139	0.139	44917	0.0 %	44917					

40
= 19
0
2
-4
1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		40	-		0 SFAM01.1	PPB	10/10/2024 15:11	Std
0.2		1145	-		0.2 SFAM01.1	PPB	10/10/2024 15:14	Std
2.5		11159	-		2.5 SFAM01.1	PPB	10/10/2024 15:17	Std
5		22658	-		5 SFAM01.1	PPB	10/10/2024 15:20	Std
7.5		32061	-		7.5 SFAM01.1	PPB	10/10/2024 15:22	Std
10		44917	-		10 SFAM01.1	PPB	10/10/2024 15:24	Std
ICV034	ICV034	16830	3.7853 -		SFAM01.1	PPB	10/10/2024 15:27	SMPL
ICB034	ICB034	-258	-0.08 -		SFAM01.1	PPB	10/10/2024 15:30	SMPL
CCV006	CCV006	22788	5.133 -		SFAM01.1	PPB	10/10/2024 15:32	SMPL
CCB006	CCB006	-258	-0.08 -		SFAM01.1	PPB	10/10/2024 15:34	SMPL
PB164039BL	PBW039	-19	-0.0259 -		SFAM01.1	PPB	10/10/2024 15:36	SMPL
P4265-01	MC0C51	-12	-0.0243 -		SFAM01.1	PPB	10/10/2024 15:39	SMPL
P4265-02	MC0C53	-10	-0.0239 -		SFAM01.1	PPB	10/10/2024 15:41	SMPL
P4265-03	MC0C69	-7	-0.0232 -		SFAM01.1	PPB	10/10/2024 15:43	SMPL
P4265-04	MC0C71	3336	0.733 -		SFAM01.1	PPB	10/10/2024 15:45	SMPL
P4265-05	MC0C73	-426	-0.118 -		SFAM01.1	PPB	10/10/2024 15:48	SMPL
P4265-06	MC0C75	21	-0.0169 -		SFAM01.1	PPB	10/10/2024 15:50	SMPL
P4265-07	MC0C77	-53	-0.0336 -		SFAM01.1	PPB	10/10/2024 15:52	SMPL
P4265-08	MC0C81	3119	0.6839 -		SFAM01.1	PPB	10/10/2024 15:55	SMPL
P4265-09	MC0C85	-128	-0.0506 -		SFAM01.1	PPB	10/10/2024 15:57	SMPL
P4265-10	MC0CA3	-192	-0.065 -		SFAM01.1	PPB	10/10/2024 15:59	SMPL
P4265-11	MC0CA5	-8	-0.0234 -		SFAM01.1	PPB	10/10/2024 16:01	SMPL
P4265-12	MC0CA5D	2	-0.0212 -		SFAM01.1	PPB	10/10/2024 16:04	SMPL
P4265-13	MC0CA5S	4443	0.9834 -		SFAM01.1	PPB	10/10/2024 16:06	SMPL
P4265-14	MC0CC3	241	0.0329 -		SFAM01.1	PPB	10/10/2024 16:08	SMPL
P4265-15	MC0CC6	28	-0.0153 -		SFAM01.1	PPB	10/10/2024 16:10	SMPL
P4265-16	MC0CD3	-67	-0.0368 -		SFAM01.1	PPB	10/10/2024 16:13	SMPL
P4265-17	MC0C47	61	-0.0078 -		SFAM01.1	PPB	10/10/2024 16:15	SMPL
P4265-18	MC0C49	-13	-0.0245 -		SFAM01.1	PPB	10/10/2024 16:17	SMPL
P4265-19	MC0C55	-45	-0.0318 -		SFAM01.1	PPB	10/10/2024 16:20	SMPL
CCV007	CCV007	22979	5.1762 -		SFAM01.1	PPB	10/10/2024 16:22	SMPL
CCB007	CCB007	-139	-0.053 -		SFAM01.1	PPB	10/10/2024 16:24	SMPL
P4265-20	MC0C57	-3	-0.0223 -		SFAM01.1	PPB	10/10/2024 16:26	SMPL
P4265-21	MC0C67	28	-0.0153 -		SFAM01.1	PPB	10/10/2024 16:29	SMPL
P4265-22	MC0C79	10	-0.0193 -		SFAM01.1	PPB	10/10/2024 16:31	SMPL
PB164040BL	40	54	-0.0094 -		SFAM01.1	PPB	10/10/2024 16:33	SMPL
P4266-01	MC0C52	57	-0.0087 -		SFAM01.1	PPB	10/10/2024 16:35	SMPL
P4266-02	MC0C54	21	-0.0169 -		SFAM01.1	PPB	10/10/2024 16:38	SMPL
P4266-03	MC0C70	59	-0.0083 -		SFAM01.1	PPB	10/10/2024 16:40	SMPL
P4266-04	MC0C72	442	0.0784 -		SFAM01.1	PPB	10/10/2024 16:42	SMPL
P4266-05	MC0C74	-98	-0.0438 -		SFAM01.1	PPB	10/10/2024 16:45	SMPL
P4266-06	MC0C76	-47	-0.0322 -		SFAM01.1	PPB	10/10/2024 16:47	SMPL
P4266-07	MC0C78	19	-0.0173 -		SFAM01.1	PPB	10/10/2024 16:49	SMPL
P4266-08	MC0C82	545	0.1017 -		SFAM01.1	PPB	10/10/2024 16:51	SMPL
P4266-09	MC0C86	-11	-0.0241 -		SFAM01.1	PPB	10/10/2024 16:54	SMPL
P4266-10	MC0CA4	-6	-0.023 -		SFAM01.1	PPB	10/10/2024 16:56	SMPL

LB132869

INSTRUMENT ID : CV1

P4266-11	MC0CA6	50	-0.0103 -	SFAM01.1	PPB	10/10/2024 16:58	SMPL
P4266-12	MC0CA6D	37	-0.0132 -	SFAM01.1	PPB	10/10/2024 17:00	SMPL
P4266-13	MC0CA6S	5267	1.1698 -	SFAM01.1	PPB	10/10/2024 17:03	SMPL
P4266-14	MC0CC4	199	0.0234 -	SFAM01.1	PPB	10/10/2024 17:05	SMPL
P4266-15	MC0CC7	8	-0.0198 -	SFAM01.1	PPB	10/10/2024 17:07	SMPL
P4266-16	MC0C48	-3	-0.0223 -	SFAM01.1	PPB	10/10/2024 17:09	SMPL
CCV008	CCV008	22702	5.1135 -	SFAM01.1	PPB	10/10/2024 17:12	SMPL
CCB008	CCB008	-97	-0.0435 -	SFAM01.1	PPB	10/10/2024 17:14	SMPL
P4266-17	MC0C50	-25	-0.0273 -	SFAM01.1	PPB	10/10/2024 17:16	SMPL
P4266-18	MC0C56	-10	-0.0239 -	SFAM01.1	PPB	10/10/2024 17:19	SMPL
P4266-19	MC0C58	1	-0.0214 -	SFAM01.1	PPB	10/10/2024 17:21	SMPL
P4266-20	MC0C68	32	-0.0144 -	SFAM01.1	PPB	10/10/2024 17:23	SMPL
P4266-21	MC0C80	-7	-0.0232 -	SFAM01.1	PPB	10/10/2024 17:25	SMPL
P4266-22	MC0C84	-9	-0.0236 -	SFAM01.1	PPB	10/10/2024 17:28	SMPL
CCV009	CCV009	22824	5.1411 -	SFAM01.1	PPB	10/10/2024 17:30	SMPL
CCB009	CCB009	-78	-0.0392 -	SFAM01.1	PPB	10/10/2024 17:32	SMPL