

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132872

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	10/10/2024	12:20
2	P4372-01		1	14	100.00	>212.0	>212.0	10/10/2024	12:50
3	P4372-01DUP		1	14	100.00	>212.0	>212.0	10/10/2024	13:20
4	P4372-02		1	12	100.00	>212.0	>212.0	10/10/2024	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06