

Analysis Method: 9034
Parameter: Sulfide
Run Number: LB132876

ANALYST: rubina
SUPERVISOR REVIEW BY: Iwona
Constant: 16000
Normality1: 0.025
Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3114
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB164037BL		1	50	50	5.00	0.00	4.94	4.94	0.06	0.00	0.00	10/10/2024	16:10
2	PB164037BS	25	1	50	50	5.00	0.00	1.92	1.92	3.08	3.02	24.16	10/10/2024	16:13
3	P4368-07		1	50	50	5.00	0.00	4.88	4.88	0.12	0.06	0.48	10/10/2024	16:15
4	P4368-08		1	50	50	5.00	0.00	4.80	4.80	0.20	0.14	1.12	10/10/2024	16:18

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume