

LB132892


SFAM01.1

1 NUTRIMENT 101 CV1

Linear

μ Abs.:

43905



A= 0.0000e+000

B= 2.2117e-004 slope

C= 3.3614e-003 y-intercept

Rho= 0.9966320

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Notes
0.00	0.000	0.018	0.018	66	0.000	66					
0.05	0.050					442	0				
0.20	0.200	0.214	0.014	954	0.0 %	954					
2.50	2.500	2.621	0.121	11837	0.0 %	11837					
5.00	5.000	4.595	-0.405	20761	0.0 %	20761					
7.50	7.500	8.037	0.537	36323	0.0 %	36323					
10.0	10.000	9.714	-0.286	43905	0.0 %	43905					

7.5)
=
7
5
- 8
7
- 3

LB132892

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	66	-		0 SFAM01.1	PPB	10/11/2024 13:51	Std
	0.2 S01	954	-		0.2 SFAM01.1	PPB	10/11/2024 13:55	Std
	2.5 S02	11837	-		2.5 SFAM01.1	PPB	10/11/2024 13:58	Std
	5 S03	20761	-		5 SFAM01.1	PPB	10/11/2024 14:00	Std
	7.5 S04	36323	-		7.5 SFAM01.1	PPB	10/11/2024 14:04	Std
	10 S05	43905	-		10 SFAM01.1	PPB	10/11/2024 14:12	Std
ICV035	ICV035	17305	3.8308 -		SFAM01.1	PPB	10/11/2024 14:15	SMPL
ICB035	ICB035	-171	-0.0345 -		SFAM01.1	PPB	10/11/2024 14:17	SMPL
CCV010	CCV010	20476	4.5321 -		SFAM01.1	PPB	10/11/2024 14:20	SMPL
CCB010	CCB010	-187	-0.038 -		SFAM01.1	PPB	10/11/2024 14:22	SMPL
PB164079BL	PBW079	37	0.0115 -		SFAM01.1	PPB	10/11/2024 14:24	SMPL
P4279-01	MCOC83	36	0.0113 -		SFAM01.1	PPB	10/11/2024 14:26	SMPL
P4279-02	MCOC87	67	0.0182 -		SFAM01.1	PPB	10/11/2024 14:29	SMPL
P4279-03	MCOCA7	-112	-0.0214 -		SFAM01.1	PPB	10/11/2024 14:31	SMPL
P4279-04	MCOCA9	59	0.0164 -		SFAM01.1	PPB	10/11/2024 14:33	SMPL
P4279-05	MCOCB3	37	0.0115 -		SFAM01.1	PPB	10/11/2024 14:35	SMPL
P4279-06	MCOCB5	-51	-0.0079 -		SFAM01.1	PPB	10/11/2024 14:38	SMPL
P4279-07	MCOCB7	-8	0.0016 -		SFAM01.1	PPB	10/11/2024 14:40	SMPL
P4279-08	MCOCB9	12	0.006 -		SFAM01.1	PPB	10/11/2024 14:42	SMPL
P4279-09	MCOCC8	-16	-0.0002 -		SFAM01.1	PPB	10/11/2024 14:45	SMPL
P4279-10	MCOCDO	64	0.0175 -		SFAM01.1	PPB	10/11/2024 14:47	SMPL
P4279-11	MCOCD4	465	0.1062 -		SFAM01.1	PPB	10/11/2024 14:49	SMPL
P4279-12	MCOCB9D	203	0.0483 -		SFAM01.1	PPB	10/11/2024 14:51	SMPL
P4279-13	MCOCB9S	4064	0.9022 -		SFAM01.1	PPB	10/11/2024 14:54	SMPL
PB164080BL	PBW080	-23	-0.0017 -		SFAM01.1	PPB	10/11/2024 14:56	SMPL
P4280-01	MCOC88	79	0.0208 -		SFAM01.1	PPB	10/11/2024 14:58	SMPL
P4280-02	MCOCA8	-82	-0.0148 -		SFAM01.1	PPB	10/11/2024 15:00	SMPL
P4280-03	MCOCB0	102	0.0259 -		SFAM01.1	PPB	10/11/2024 15:03	SMPL
P4280-04	MCOCB4	149	0.0363 -		SFAM01.1	PPB	10/11/2024 15:05	SMPL
P4280-05	MCOCB6	51	0.0146 -		SFAM01.1	PPB	10/11/2024 15:07	SMPL
CCV011	CCV011	20223	4.4762 -		SFAM01.1	PPB	10/11/2024 15:09	SMPL
CCB011	CCB011	-169	-0.034 -		SFAM01.1	PPB	10/11/2024 15:12	SMPL
P4280-06	MCOCB8	38	0.0118 -		SFAM01.1	PPB	10/11/2024 15:14	SMPL
P4280-07	MCOCC0	56	0.0157 -		SFAM01.1	PPB	10/11/2024 15:16	SMPL
P4280-08	MCOCC9	30	0.01 -		SFAM01.1	PPB	10/11/2024 15:19	SMPL
P4280-09	MCOCD1	57	0.016 -		SFAM01.1	PPB	10/11/2024 15:21	SMPL
P4280-10	MCOCD1D	134	0.033 -		SFAM01.1	PPB	10/11/2024 15:23	SMPL
P4280-11	MCOCD1S	4477	0.9936 -		SFAM01.1	PPB	10/11/2024 15:25	SMPL
CCV012	CCV012	20409	4.5173 -		SFAM01.1	PPB	10/11/2024 15:28	SMPL
CCB012	CCB012	-150	-0.0298 -		SFAM01.1	PPB	10/11/2024 15:30	SMPL