

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132897

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/11/2024	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/11/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/11/2024	16:29
4	ICV	1	Water	NA	NA	20.3	7.01	10/11/2024	16:30
5	CCV1	1	Water	NA	NA	20.2	2.01	10/11/2024	16:35
6	P4396-01	1	Water	NA	NA	20.1	7.65	10/11/2024	16:40
7	P4396-01DUP	1	Water	NA	NA	20.2	7.66	10/11/2024	16:41
8	CCV2	1	Water	NA	NA	20.3	12.02	10/11/2024	16:45

WORKLIST(Hardcopy Internal Chain)

WB 132897

WorkList Name : PH P396 WorkList ID : 184375 Department : Wet-Chemistry Date : 10-11-2024 16:12:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4396-01	WASTE-WATER-FRAC-TANK	Water	pH	Cool 4 deg C	PSEG03	K32	10/11/2024	9040C

Date/Time 10-11-24 16:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10-11-24 17:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]