

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132932

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/14/2024	15:45
2	P4396-01		1	12	100.00	>212.0	>212.0	10/14/2024	16:15
3	P4396-01DUP		1	12	100.00	>212.0	>212.0	10/14/2024	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132932

WorkList Name : FP--1014

WorkList ID : 184405

Department : Wet-Chemistry

Date : 10-14-2024 09:28:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4396-01	WASTE-WATER-FRAC-TANK	Water	Flash Point	Cool 4 deg C	PSEG03	K32	10/11/2024	1010B

Date/Time 10/14/2024 15:35
Raw Sample Received by: RM
Raw Sample Relinquished by: JH 10/14/2024

Date/Time 10/14/2024 17:25
Raw Sample Received by: JH 10/14/2024
Raw Sample Relinquished by: RM 10/14/2024