

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132933

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	10/15/2024	08:20
2	P4402-01		1	12	100.00	>212.0	>212.0	10/15/2024	08:50
3	P4402-01DUP		1	12	100.00	>212.0	>212.0	10/15/2024	09:20
4	P4402-02		1	14	100.00	>212.0	>212.0	10/15/2024	09:50
5	P4402-03		1	11	100.00	>212.0	>212.0	10/15/2024	10:20
6	P4402-04		1	12	100.00	>212.0	>212.0	10/15/2024	10:50
7	P4402-05		1	14	100.00	>212.0	>212.0	10/15/2024	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132933

WorkList Name : fp-10-24*

WorkList ID : 184439

Department : Wet-Chemistry

Date : 10-14-2024 17:36:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4402-01 D	1 - Drum	Water	Flash Point	Cool 4 deg C	PSEG03	K11	10/14/2024	1010B
P4402-02 D	2 - Drum	Water	Flash Point	Cool 4 deg C	PSEG03	K11	10/14/2024	1010B
P4402-03 D	3 - Tote	Water	Flash Point	Cool 4 deg C	PSEG03	K11	10/14/2024	1010B
P4402-04 D	4 - Tote	Water	Flash Point	Cool 4 deg C	PSEG03	K11	10/14/2024	1010B
P4402-05 D	5 - Tote	Water	Flash Point	Cool 4 deg C	PSEG03	K11	10/14/2024	1010B

Date/Time 10/15/2024 08:10
Raw Sample Received by: RM
Raw Sample Relinquished by: M. J. 1

Date/Time 10/15/2024 12:00
Raw Sample Received by: M. J. 602C
Raw Sample Relinquished by: RM 602C