


LB132940

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	210	0.000	210	0				-
0.2	0.200	0.184	-0.016	922	0.0 %	922					-8
2.5	2.500	2.396	-0.104	10694	0.0 %	10694					-4
5.0	5.000	5.080	0.080	22550	0.0 %	22550					2
7.5	7.500	7.640	0.140	33860	0.0 %	33860					2
10.0	10.000	9.876	-0.124	43739	0.0 %	43739					1

LB132940 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Oct 2024 11:50:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	210	-	0.0000	7470A	PPB	15 Oct 2024 12:15:03	S	Std
0.2 - 1	50.2	922	-	0.2000	7470A	PPB	15 Oct 2024 12:17:20	S	Std
2.5 - 1	52.5	10694	-	2.5000	7470A	PPB	15 Oct 2024 12:19:37	S	Std
5.0 - 1	55	22550	-	5.0000	7470A	PPB	15 Oct 2024 12:21:54	S	Std
7.5 - 1	57.5	33860	-	7.5000	7470A	PPB	15 Oct 2024 12:24:10	S	Std
10.0 - 1	510	43739	-	10.0000	7470A	PPB	15 Oct 2024 12:29:07	S	Std
ICV37 - 1	ICV37	17501	3.9371	-	7470A	PPB	15 Oct 2024 12:31:59	U	SMPL
ICB37 - 1	ICB37	-25	-0.0299	-	7470A	PPB	15 Oct 2024 12:34:14	U	SMPL
CCV21 - 1	CCV21	22865	5.1512	-	7470A	PPB	15 Oct 2024 12:36:30	U	SMPL
CCB21 - 1	CCB21	-9	-0.0263	-	7470A	PPB	15 Oct 2024 12:38:46	U	SMPL
CRA - 1	CRA	911	0.1820	-	7470A	PPB	15 Oct 2024 12:41:03	U	SMPL
HighStd - 1	HighStd	43214	9.7572	-	7470A	PPB	15 Oct 2024 12:43:18	U	SMPL
ChkStd - 1	ChkStd	30574	6.8961	-	7470A	PPB	15 Oct 2024 12:45:33	U	SMPL
PB164159BL - 1	PBW	-322	-0.0971	-	7470A	PPB	15 Oct 2024 12:47:51	U	SMPL
PB164159BS - 1	LCSW	17291	3.8896	-	7470A	PPB	15 Oct 2024 12:52:45	U	SMPL
P4375-02 - 1	CONCRETE-PADS-1	-197	-0.0688	-	7470A	PPB	15 Oct 2024 12:55:01	U	SMPL
P4375-04 - 1	CONCRETE-PADS-2	-18	-0.0283	-	7470A	PPB	15 Oct 2024 12:57:19	U	SMPL
P4375-06 - 1	CONCRETE-PADS-3	6	-0.0229	-	7470A	PPB	15 Oct 2024 12:59:35	U	SMPL
P4375-08 - 1	CONCRETE-PADS-4	0	-0.0242	-	7470A	PPB	15 Oct 2024 13:01:53	U	SMPL
P4375-10 - 1	CONCRETE-PADS-5	-8	-0.0260	-	7470A	PPB	15 Oct 2024 13:04:08	U	SMPL
CCV22 - 1	CCV22	22759	5.1272	-	7470A	PPB	15 Oct 2024 13:06:23	U	SMPL
CCB22 - 1	CCB22	-140	-0.0559	-	7470A	PPB	15 Oct 2024 13:08:38	U	SMPL
P4375-12 - 1	CONCRETE-PADS-6	-31	-0.0312	-	7470A	PPB	15 Oct 2024 13:10:55	U	SMPL
P4375-14 - 1	CONCRETE-PADS-7	31	-0.0172	-	7470A	PPB	15 Oct 2024 13:13:10	U	SMPL
P4375-16 - 1	CONCRETE-PADS-8	80	-0.0061	-	7470A	PPB	15 Oct 2024 13:15:25	U	SMPL
P4375-18 - 1	CONCRETE-PADS-9	56	-0.0116	-	7470A	PPB	15 Oct 2024 13:17:40	U	SMPL
P4375-20 - 1	CONCRETE-PADS-10	13	-0.0213	-	7470A	PPB	15 Oct 2024 13:19:56	U	SMPL
P4375-22 - 1	CONCRETE-PADS-11	0	-0.0242	-	7470A	PPB	15 Oct 2024 13:22:11	U	SMPL
P4375-24 - 1	FENCE-POST-1	45	-0.0140	-	7470A	PPB	15 Oct 2024 13:24:27	U	SMPL
P4375-26 - 1	FENCE-POST-2	0	-0.0242	-	7470A	PPB	15 Oct 2024 13:26:44	U	SMPL
P4375-28 - 1	FENCE-POST-3	35	-0.0163	-	7470A	PPB	15 Oct 2024 13:29:01	U	SMPL
P4375-28DUP - 1	FENCE-POST-3DUP	89	-0.0041	-	7470A	PPB	15 Oct 2024 13:31:17	U	SMPL
CCV23 - 1	CCV23	21910	4.9351	-	7470A	PPB	15 Oct 2024 13:33:32	U	SMPL
CCB23 - 1	CCB23	8	-0.0224	-	7470A	PPB	15 Oct 2024 13:35:47	U	SMPL
P4375-28MS - 1	FENCE-POST-3MS	15630	3.5136	-	7470A	PPB	15 Oct 2024 13:38:04	U	SMPL
P4375-28MSD - 1	FENCE-POST-3MSD	17185	3.8656	-	7470A	PPB	15 Oct 2024 13:40:20	U	SMPL
P4395-02 - 1	F05308-SOLID	171	0.0145	-	7470A	PPB	15 Oct 2024 13:42:37	U	SMPL
P4403-02 - 1	Hawthorne TP Soil	400	0.0663	-	7470A	PPB	15 Oct 2024 13:44:54	U	SMPL
PB164160BL - 1	PBW	27	-0.0181	-	7470A	PPB	15 Oct 2024 13:47:10	U	SMPL
PB164160BS - 1	LCSW	18173	4.0892	-	7470A	PPB	15 Oct 2024 13:49:26	U	SMPL
P4354-01 - 1	291551	-73	-0.0408	-	7470A	PPB	15 Oct 2024 13:51:42	U	SMPL
P4354-01DUP - 1	291551DUP	125	0.0041	-	7470A	PPB	15 Oct 2024 13:54:01	U	SMPL
P4354-01MS - 1	291551MS	15915	3.5781	-	7470A	PPB	15 Oct 2024 13:56:17	U	SMPL
P4354-01MSD - 1	291551MSD	15950	3.5860	-	7470A	PPB	15 Oct 2024 13:58:33	U	SMPL
CCV24 - 1	CCV24	21657	4.8778	-	7470A	PPB	15 Oct 2024 14:00:51	U	SMPL
CCB24 - 1	CCB24	-214	-0.0727	-	7470A	PPB	15 Oct 2024 14:03:09	U	SMPL
P4370-02 - 1	COMP	855	0.1693	-	7470A	PPB	15 Oct 2024 14:05:28	U	SMPL
P4382-01 - 1	DECON-DRUM	1116	0.2284	-	7470A	PPB	15 Oct 2024 14:07:43	U	SMPL
P4396-01 - 1	WASTE-WATER-FRAC-TANK	518	0.0930	-	7470A	PPB	15 Oct 2024 14:09:59	U	SMPL
P4397-04 - 1	WB-301-SW	545	0.0991	-	7470A	PPB	15 Oct 2024 14:12:17	U	SMPL
PB164021TB - 1	PB164021TB	-52	-0.0360	-	7470A	PPB	15 Oct 2024 14:14:33	U	SMPL
PB164096TB - 1	PB164096TB	1	-0.0240	-	7470A	PPB	15 Oct 2024 14:16:49	U	SMPL
P4375-28LX5 - 1		-52	-0.0360	-	7470A	PPB	15 Oct 2024 14:19:05	U	SMPL
P4375-28A - 1		16393	3.6863	-	7470A	PPB	15 Oct 2024 14:21:22	U	SMPL
P4354-01LX5 - 1		-209	-0.0715	-	7470A	PPB	15 Oct 2024 14:23:39	U	SMPL
P4354-01A - 1		16611	3.7356	-	7470A	PPB	15 Oct 2024 14:25:56	U	SMPL
CCV25 - 1	CCV25	20959	4.7198	-	7470A	PPB	15 Oct 2024 14:28:12	U	SMPL
CCB25 - 1	CCB25	-112	-0.0496	-	7470A	PPB	15 Oct 2024 14:30:29	U	SMPL