

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132953

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/15/2024	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	10/15/2024	16:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/15/2024	16:14
4	ICV	1	Water	NA	NA	20.3	7.02	10/15/2024	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	10/15/2024	16:22
6	P4405-01	1	Water	NA	NA	20.4	7.13	10/15/2024	16:25
7	P4405-01DUP	1	Water	NA	NA	20.5	7.14	10/15/2024	16:26
8	P4407-01	1	Water	NA	NA	20.8	5.85	10/15/2024	16:33
9	CCV2	1	Water	NA	NA	20.3	12.02	10/15/2024	16:37