

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132955

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/16/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.01	10/16/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	10/16/2024	08:17
4	ICV	1	Water	NA	NA	20.3	7.00	10/16/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	10/16/2024	08:25
6	P4408-01	1	Solid	20.02	20	22.1	5.60	10/16/2024	08:30
7	P4408-01DUP	1	Solid	20.03	20	22.2	5.61	10/16/2024	08:31
8	CCV2	1	Water	NA	NA	20.2	12.02	10/16/2024	08:35

WORKLIST(Hardcopy Internal Chain)

W 132955

WorkList Name : ph p4408 s WorkList ID : 184474 Department : Wet-Chemistry Date : 10-16-2024 07:50:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4408-01	1363	Solid	pH	Cool 4 deg C	PSEG03	K41	10/15/2024	9045D

Date/Time 10/16/24 08:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/16/24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]