

LB132956


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4159E



A= 0.0000e+000

B= 2.4232e-004 slope

C= -9.3447e-003 y-intercept

Rho= 0.9994796

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/dl
0.00	0.000	0.029	0.029	160	0.000	160					=
0.05	0.050					442	0				16
0.20	0.200	0.231	0.031	991	0.0 %	991					-5
2.50	2.500	2.377	-0.123	9846	0.0 %	9846					3
5.00	5.000	5.173	0.173	21388	0.0 %	21388					2
7.50	7.500	7.319	-0.181	30241	0.0 %	30241					-
10.0	10.000	10.071	0.071	41598	0.0 %	41598					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	160	-		0 SFAM01.1	PPB	10/16/2024 9:10	Std
	0.2 S01	991	-		0.2 SFAM01.1	PPB	10/16/2024 9:12	Std
	2.5 S02	9846	-		2.5 SFAM01.1	PPB	10/16/2024 9:15	Std
	5 S03	21388	-		5 SFAM01.1	PPB	10/16/2024 9:17	Std
	7.5 S04	30241	-		7.5 SFAM01.1	PPB	10/16/2024 9:19	Std
	10 S05	41598	-		10 SFAM01.1	PPB	10/16/2024 9:24	Std
ICV038	ICV038	16475	3.983 -		SFAM01.1	PPB	10/16/2024 9:28	SMPL
ICB038	ICB038	-113	-0.0367 -		SFAM01.1	PPB	10/16/2024 9:30	SMPL
CCV026	CCV026	21528	5.2074 -		SFAM01.1	PPB	10/16/2024 9:33	SMPL
CCB026	CCB026	-172	-0.051 -		SFAM01.1	PPB	10/16/2024 9:35	SMPL
PB164184BL	PBW184	86	0.0115 -		SFAM01.1	PPB	10/16/2024 9:37	SMPL
P4319-01	MC0C59	100	0.0149 -		SFAM01.1	PPB	10/16/2024 9:39	SMPL
P4319-02	MC0C61	102	0.0154 -		SFAM01.1	PPB	10/16/2024 9:42	SMPL
P4319-03	MC0C63	136	0.0236 -		SFAM01.1	PPB	10/16/2024 9:44	SMPL
P4319-04	MC0C65	96	0.0139 -		SFAM01.1	PPB	10/16/2024 9:46	SMPL
P4319-05	MC0C89	153	0.0277 -		SFAM01.1	PPB	10/16/2024 9:49	SMPL
P4319-06	MC0C91	1316	0.3096 -		SFAM01.1	PPB	10/16/2024 9:51	SMPL
P4319-07	MC0C93	3490	0.8364 -		SFAM01.1	PPB	10/16/2024 9:53	SMPL
P4319-08	MC0C95	654	0.1491 -		SFAM01.1	PPB	10/16/2024 9:55	SMPL
P4319-09	MC0C97	254	0.0522 -		SFAM01.1	PPB	10/16/2024 9:58	SMPL
P4319-10	MC0C99	76	0.0091 -		SFAM01.1	PPB	10/16/2024 10:00	SMPL
P4319-11	MC0CA1	179	0.034 -		SFAM01.1	PPB	10/16/2024 10:02	SMPL
P4319-12	MC0CB1	169	0.0316 -		SFAM01.1	PPB	10/16/2024 10:05	SMPL
P4319-13	MC0CC1	220	0.044 -		SFAM01.1	PPB	10/16/2024 10:07	SMPL
P4319-14	MC0CC1D	163	0.0302 -		SFAM01.1	PPB	10/16/2024 10:09	SMPL
P4319-15	MC0CC1S	4578	1.1 -		SFAM01.1	PPB	10/16/2024 10:11	SMPL
P4319-16	MC0CD5	101	0.0151 -		SFAM01.1	PPB	10/16/2024 10:14	SMPL
PB164185BL	PBW185	143	0.0253 -		SFAM01.1	PPB	10/16/2024 10:16	SMPL
P4240-01	MH2B87	109	0.0171 -		SFAM01.1	PPB	10/16/2024 10:18	SMPL
P4240-02	MH2B88	110	0.0173 -		SFAM01.1	PPB	10/16/2024 10:20	SMPL
CCV027	CCV027	21093	5.102 -		SFAM01.1	PPB	10/16/2024 10:23	SMPL
CCB027	CCB027	23	-0.0038 -		SFAM01.1	PPB	10/16/2024 10:25	SMPL
P4240-03	MH2B88D	89	0.0122 -		SFAM01.1	PPB	10/16/2024 10:27	SMPL
P4240-04	MH2B88S	3972	0.9532 -		SFAM01.1	PPB	10/16/2024 10:29	SMPL
P4320-01	MC0C60	70	0.0076 -		SFAM01.1	PPB	10/16/2024 10:32	SMPL
P4320-02	MC0C62	121	0.02 -		SFAM01.1	PPB	10/16/2024 10:34	SMPL
P4320-03	MC0C64	125	0.0209 -		SFAM01.1	PPB	10/16/2024 10:36	SMPL
P4320-04	MC0C66	103	0.0156 -		SFAM01.1	PPB	10/16/2024 10:38	SMPL
P4320-05	MC0C90	164	0.0304 -		SFAM01.1	PPB	10/16/2024 10:41	SMPL
P4320-06	MC0C92	257	0.0529 -		SFAM01.1	PPB	10/16/2024 10:43	SMPL
P4320-07	MC0C94	1165	0.273 -		SFAM01.1	PPB	10/16/2024 10:45	SMPL
P4320-08	MC0C96	509	0.114 -		SFAM01.1	PPB	10/16/2024 10:48	SMPL
P4320-09	MC0C98	185	0.0355 -		SFAM01.1	PPB	10/16/2024 10:50	SMPL
P4320-10	MC0CA0	100	0.0149 -		SFAM01.1	PPB	10/16/2024 10:52	SMPL
P4320-11	MC0CA2	88	0.012 -		SFAM01.1	PPB	10/16/2024 10:54	SMPL
P4320-12	MC0CB2	271	0.0563 -		SFAM01.1	PPB	10/16/2024 10:57	SMPL

LB132956

INSTRUMENT ID : CV1

P4320-13	MC0CC2	67	0.0069 -	SFAM01.1	PPB	10/16/2024 10:59 SMPL
P4320-14	MC0CC2D	79	0.0098 -	SFAM01.1	PPB	10/16/2024 11:01 SMPL
P4320-15	MC0CC2S	3966	0.9517 -	SFAM01.1	PPB	10/16/2024 11:06 SMPL
CCV028	CCV028	21979	5.3167 -	SFAM01.1	PPB	10/16/2024 11:08 SMPL
CCB028	CCB028	9	-0.0072 -	SFAM01.1	PPB	10/16/2024 11:10 SMPL