

7471B

INSTRUMENT 11) CV 1

Linear

 $\mu \text{ Abs} :$

43641

A= 0.0000e+000

B= 2.3085e-004

C= -2.6234e-002

Rho= 0.9999518

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	-0.006	-0.006	86	0.000	86					-
0.2	0.200	0.238	0.038	1144	0.0 %	1144					19
2.5	2.500	2.479	-0.021	10854	0.0 %	10854					-1
5.0	5.000	5.001	0.001	21778	0.0 %	21778					0
7.5	7.500	7.440	-0.060	32341	0.0 %	32341					
10.0	10.000	10.048	0.048	43641	0.0 %	43641					

LB132960 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Oct 2024 12:42:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.000	7471B	PPB	16 Oct 2024 13:33:01	S	Std
0.2 - 1	50-2	1144	-	0.200	7471B	PPB	16 Oct 2024 13:35:18	S	Std
2.5 - 1	52-5	10854	-	2.500	7471B	PPB	16 Oct 2024 13:37:34	S	Std
5.0 - 1	55	21778	-	5.000	7471B	PPB	16 Oct 2024 13:39:51	S	Std
7.5 - 1	57-5	32341	-	7.500	7471B	PPB	16 Oct 2024 13:47:43	S	Std
10.0 - 1	57-5	43641	-	10.000	7471B	PPB	16 Oct 2024 13:49:59	S	Std
ICV39 - 1	ICV39	17805	4.0840	-	7471B	PPB	16 Oct 2024 13:53:22	U	SMPL
ICB39 - 1	ICB39	-296	-0.0946	-	7471B	PPB	16 Oct 2024 13:55:37	U	SMPL
CCV29 - 1	CCV29	22285	5.1182	-	7471B	PPB	16 Oct 2024 13:57:54	U	SMPL
CCB29 - 1	CCB29	-279	-0.0906	-	7471B	PPB	16 Oct 2024 14:00:10	U	SMPL
CRA - 1	CRA	905	0.1827	-	7471B	PPB	16 Oct 2024 14:05:31	U	SMPL
HighStd - 1	HighStd	43367	9.9850	-	7471B	PPB	16 Oct 2024 14:07:46	U	SMPL
ChkStd - 1	ChkStd	29643	6.8168	-	7471B	PPB	16 Oct 2024 14:10:01	U	SMPL
PB164194BL - 1	PBS	39	-0.0172	-	7471B	PPB	16 Oct 2024 14:15:08	U	SMPL
PB164194BS - 1	LCSS	17415	3.9940	-	7471B	PPB	16 Oct 2024 14:17:23	U	SMPL
P4397-01 - 1	WB-301-TOP	250411	57.7810	-	7471B	PPB	16 Oct 2024 14:19:39	U	SMPL
P4397-02 - 1	WB-301-BOT	1845	0.3997	-	7471B	PPB	16 Oct 2024 14:21:58	U	SMPL
P4400-01 - 1	NB-08-101424	4303	0.9671	-	7471B	PPB	16 Oct 2024 14:24:44	U	SMPL
P4401-01 - 1	SU-03-101424	585	0.1088	-	7471B	PPB	16 Oct 2024 14:27:02	U	SMPL
P4403-01 - 1	Hawthorne TP Soil	1138	0.2365	-	7471B	PPB	16 Oct 2024 14:29:20	U	SMPL
CCV30 - 1	CCV30	22686	5.2108	-	7471B	PPB	16 Oct 2024 14:31:35	U	SMPL
CCB30 - 1	CCB30	-205	-0.0736	-	7471B	PPB	16 Oct 2024 14:33:50	U	SMPL
P4406-01 - 1	OK-02-101524	2882	0.6391	-	7471B	PPB	16 Oct 2024 14:36:08	U	SMPL
P4410-01 - 1	TR-04-101524	1277	0.2686	-	7471B	PPB	16 Oct 2024 14:38:23	U	SMPL
P4410-01DUP - 1	TR-04-101524DUP	1123	0.2330	-	7471B	PPB	16 Oct 2024 14:40:40	U	SMPL
P4410-01MS - 1	TR-04-101524MS	21265	4.8828	-	7471B	PPB	16 Oct 2024 14:42:56	U	SMPL
P4410-01MSD - 1	TR-04-101524MSD	21871	5.0227	-	7471B	PPB	16 Oct 2024 14:45:11	U	SMPL
P4410-01LX5 - 1		88	-0.0059	-	7471B	PPB	16 Oct 2024 14:47:30	U	SMPL
P4410-01A - 1		18271	4.1916	-	7471B	PPB	16 Oct 2024 14:49:49	U	SMPL
P4397-01DLX10 - 1		26460	6.0820	-	7471B	PPB	16 Oct 2024 14:52:06	U	SMPL
CCV31 - 1	CCV31	22839	5.2461	-	7471B	PPB	16 Oct 2024 14:54:26	U	SMPL
CCB31 - 1	CCB31	-176	-0.0669	-	7471B	PPB	16 Oct 2024 14:56:44	U	SMPL