

TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 10/16/2024 14:30 **TEMP1 OUT:** 103 °C 10/17/2024 07:30
TEMP2 IN: 104 °C 10/17/2024 08:00 **TEMP2 OUT:** 104 °C 10/17/2024 09:30
TEMP3 IN: 540 °C 10/17/2024 10:00 **TEMP3 OUT:** 540 °C 10/17/2024 11:30
TEMP4 IN: 560 °C 10/17/2024 12:00 **TEMP4 OUT:** 560 °C 10/17/2024 13:30

Run Number: LB132968

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB132968BL	85.9631	85.9631	100	85.9631	85.9631	85.9631	85.9631	0.0000	0
2	P4405-01	90.2644	90.2644	100	92.5578	92.5578	91.9977	91.9977	0.5601	5601
3	P4405-01DUP	85.0560	85.0560	100	87.3495	87.3495	86.7730	86.7730	0.5765	5765

A = Sample Volume (ml)
B = Final Dish + Samplewt Drying @550 (±50) °C (g)
C = Final Dish + SampleWt Drying @103-@105 °C (g)
D = Weight (g)
E = Final Empty Dish Weight (g)
F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = **C - B**

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

JB 122968

WorkList Name : tvs water p4405 WorkList ID : 184496 Department : Wet-Chemistry Date : 10-16-2024 11:00:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4405-01	MH-121	Water	TVS	Cool 4 deg C	PSEG03	K41	10/15/2024	SM2540 G

Date/Time 10.16.24 12:00
Raw Sample Received by: JB WLC
Raw Sample Relinquished by: JB SM

Date/Time 10.16.24 15:00
Raw Sample Received by: JB SM
Raw Sample Relinquished by: JB WLC