

LB132974


7471B

INSTRUMENT ID: CV4

Linear

μ Abs.:

47303



A= 0.0000e+000

B= 2.0970e-004

C= -8.3189e-003

Rho= 0.9998602

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.003	-0.003	23	0.000	23					-
0.2	0.200	0.194	-0.006	967	0.0 %	967					-3
2.5	2.500	2.465	-0.035	11795	0.0 %	11795					-1
5.0	5.000	5.017	0.017	23963	0.0 %	23963					0
7.5	7.500	7.616	0.116	36357	0.0 %	36357					2
10.0	10.000	9.911	-0.089	47303	0.0 %	47303					-1

LB132974 INSTRUMENT ID: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Oct 2024 10:51:20

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	23	-	0.0000	7471B	PPB	17 Oct 2024 10:51:36	S	Std
0.2 - 1	SO2	967	-	0.2000	7471B	PPB	17 Oct 2024 10:53:53	S	Std
2.5 - 1	SO2.5	11795	-	2.5000	7471B	PPB	17 Oct 2024 10:56:09	S	Std
5.0 - 1	SS	23963	-	5.0000	7471B	PPB	17 Oct 2024 10:58:26	S	Std
7.5 - 1	SS7.5	36357	-	7.5000	7471B	PPB	17 Oct 2024 11:00:44	S	Std
10.0 - 1	SS10	47303	-	10.0000	7471B	PPB	17 Oct 2024 11:03:03	S	Std
ICV40 - 1	ICV40	18552	3.8821	-	7471B	PPB	17 Oct 2024 11:05:56	U	SMPL
ICB40 - 1	ICB40	-304	-0.0721	-	7471B	PPB	17 Oct 2024 11:08:13	U	SMPL
CCV32 - 1	CCV32	23941	5.0122	-	7471B	PPB	17 Oct 2024 11:10:31	U	SMPL
CCB32 - 1	CCB32	-346	-0.0809	-	7471B	PPB	17 Oct 2024 11:12:45	U	SMPL
CRA - 1	CRA	908	0.1821	-	7471B	PPB	17 Oct 2024 11:15:04	U	SMPL
HighStd - 1	HighStd	46957	9.8387	-	7471B	PPB	17 Oct 2024 11:17:19	U	SMPL
ChkStd - 1	ChkStd	33061	6.9247	-	7471B	PPB	17 Oct 2024 11:19:35	U	SMPL
PB164217BL - 1	PBS	-55	-0.0199	-	7471B	PPB	17 Oct 2024 11:24:37	U	SMPL
PB164217BS - 1	LCSS	18121	3.7917	-	7471B	PPB	17 Oct 2024 11:26:52	U	SMPL
P2403-03 - 1	MDL-SOIL-03-QT2-2024	235	0.0410	-	7471B	PPB	17 Oct 2024 11:34:17	U	SMPL
P4417-02 - 1	HOPPER-A-SOLID	156	0.0244	-	7471B	PPB	17 Oct 2024 11:36:33	U	SMPL
P4417-05 - 1	HOPPER-B-SOLID	-291	-0.0693	-	7471B	PPB	17 Oct 2024 11:38:49	U	SMPL
P4417-08 - 1	HOPPER-C-SOLID	1268	0.2576	-	7471B	PPB	17 Oct 2024 11:41:07	U	SMPL
P4419-01 - 1	IB-1	73659	15.4382	-	7471B	PPB	17 Oct 2024 11:43:22	U	SMPL
CCV33 - 1	CCV33	23872	4.9977	-	7471B	PPB	17 Oct 2024 11:45:39	U	SMPL
CCB33 - 1	CCB33	-375	-0.0870	-	7471B	PPB	17 Oct 2024 11:48:04	U	SMPL
P4421-01 - 1	EO-02-101624	7886	1.6454	-	7471B	PPB	17 Oct 2024 11:50:22	U	SMPL
P4422-01 - 1	EO-01-101624	3043	0.6298	-	7471B	PPB	17 Oct 2024 11:52:38	U	SMPL
P4422-01DUP - 1	EO-01-101624DUP	2935	0.6072	-	7471B	PPB	17 Oct 2024 11:54:56	U	SMPL
P4422-01MS - 1	EO-01-101624MS	25646	5.3697	-	7471B	PPB	17 Oct 2024 11:57:12	U	SMPL
P4422-01MSD - 1	EO-01-101624MSD	24823	5.1971	-	7471B	PPB	17 Oct 2024 11:59:30	U	SMPL
P2403-01 - 1	LOD-MDL-SOIL	209	0.0355	-	7471B	PPB	17 Oct 2024 12:04:28	U	SMPL
P2403-01RE - 1	LOD-MDL-SOILRE	652	0.1284	-	7471B	PPB	17 Oct 2024 12:06:44	U	SMPL
P4422-01LX5 - 1		568	0.1108	-	7471B	PPB	17 Oct 2024 12:09:01	U	SMPL
P4422-01A - 1		23081	4.8318	-	7471B	PPB	17 Oct 2024 12:11:18	U	SMPL
P4419-01DLX5 - 1		16017	3.3505	-	7471B	PPB	17 Oct 2024 12:13:33	U	SMPL
CCV34 - 1	CCV34	24943	5.2223	-	7471B	PPB	17 Oct 2024 12:15:51	U	SMPL
CCB34 - 1	CCB34	-322	-0.0758	-	7471B	PPB	17 Oct 2024 12:18:09	U	SMPL