

Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB132979
Thermometer ID: Flash Point

Reviewed By: rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 776.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	10/17/2024	09:15
2	P4405-01		1	12	100.00	>212.0	>212.0	10/17/2024	09:45
3	P4405-01DUP		1	12	100.00	>212.0	>212.0	10/17/2024	10:15
4	P4407-01		1	11	100.00	>212.0	>212.0	10/17/2024	10:45
5	P4417-10		1	14	100.00	>212.0	>212.0	10/17/2024	11:15
6	P4417-12		1	12	100.00	>212.0	>212.0	10/17/2024	11:45
7	P4420-01		1	7	67.00	152.6	151.6	10/17/2024	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB132979

WorkList Name : FP-10-16 WorkList ID : 184512 Department : Wet-Chemistry Date : 10-16-2024 16:43:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4405-01	MH-121	Water	Flash Point	Cool 4 deg C	PSEG03	K41	10/15/2024	1010B
P4407-01	MR-CAM-06	Water	Flash Point	Cool 4 deg C	PSEG03	K41	10/15/2024	1010B
P4417-10	HOPPER-A-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	K41	10/16/2024	1010B
P4417-12	HOPPER-B-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	K41	10/16/2024	1010B
P4420-01	HCC1	Water	Flash Point	Cool 4 deg C	PSEG03	K51	10/16/2024	1010B

Date/Time 10/17/2024 09:05
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH CWC

Date/Time 10/17/2024 12:55
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: RM CWC