

LB 132982

245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37001

A= 0.0000e+000

B= 2.7080e-004

C= -1.4172e-002

Rho= 0.9999650

Accept=Accepted

Max. Conc.: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	30	0.000	30					1
0.05	0.050	0.051	0.001	240	0.0 %	240	0				2
0.2	0.200	0.172	-0.028	687	0.0 %	687					-14
2.5	2.500	2.528	0.028	9387	0.0 %	9387					1
5.0	5.000	5.051	0.051	18703	0.0 %	18703					1
7.5	7.500	7.449	-0.051	27561	0.0 %	27561					1
10.0	10.000	10.006	0.006	37001	0.0 %	37001					0

LB132982 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Oct 2024 13:58:49

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	30	-	0.0000	245.1	PPB	17 Oct 2024 14:06:38	S	Std
0.05 - 1	50.05	240	-	0.0500	245.1	PPB	17 Oct 2024 14:08:55	S	Std
0.2 - 1	50.2	687	-	0.2000	245.1	PPB	17 Oct 2024 14:11:12	S	Std
2.5 - 1	52.5	9387	-	2.5000	245.1	PPB	17 Oct 2024 14:13:29	S	Std
5.0 - 1	55	18703	-	5.0000	245.1	PPB	17 Oct 2024 14:15:46	S	Std
7.5 - 1	57.5	27561	-	7.5000	245.1	PPB	17 Oct 2024 14:18:03	S	Std
10.0 - 1	59	37001	-	10.0000	245.1	PPB	17 Oct 2024 14:28:48	S	Std
ICV41 - 1	ICV41	14728	3.9742	-	245.1	PPB	17 Oct 2024 14:36:57	U	SMPL
ICB41 - 1	ICB41	-138	-0.0515	-	245.1	PPB	17 Oct 2024 14:39:11	U	SMPL
CCV35 - 1	CCV35	19094	5.1565	-	245.1	PPB	17 Oct 2024 14:41:27	U	SMPL
CCB35 - 1	CCB35	-163	-0.0583	-	245.1	PPB	17 Oct 2024 14:43:42	U	SMPL
CRA - 1	CRA	240	0.0508	-	245.1	PPB	17 Oct 2024 14:45:59	U	SMPL
HighStd - 1	HighStd	37665	10.1855	-	245.1	PPB	17 Oct 2024 14:48:15	U	SMPL
ChkStd - 1	ChkStd	26152	7.0678	-	245.1	PPB	17 Oct 2024 14:50:29	U	SMPL
PB164225BL - 1	PBW	-33	-0.0231	-	245.1	PPB	17 Oct 2024 14:55:30	U	SMPL
PB164225BS - 1	LCSW	14528	3.9200	-	245.1	PPB	17 Oct 2024 14:57:46	U	SMPL
P4347-01 - 1	EFFLUENT-DAY-1-METALS	816	0.2068	-	245.1	PPB	17 Oct 2024 15:02:20	U	SMPL
P4347-01DUP - 1	EFFLUENT-DAY-1-METALSDUP	179	0.0343	-	245.1	PPB	17 Oct 2024 15:04:37	U	SMPL
P4347-01MS - 1	EFFLUENT-DAY-1-METALSMS	14308	3.8604	-	245.1	PPB	17 Oct 2024 15:06:54	U	SMPL
P4347-01MSD - 1	EFFLUENT-DAY-1-METALMSD	12680	3.4196	-	245.1	PPB	17 Oct 2024 15:09:09	U	SMPL
CCV36 - 1	CCV36	17771	4.7982	-	245.1	PPB	17 Oct 2024 15:11:25	U	SMPL
CCB36 - 1	CCB36	-180	-0.0629	-	245.1	PPB	17 Oct 2024 15:13:42	U	SMPL
P2403-10 - 1	LOD-MDL-WATER	255	0.0549	-	245.1	PPB	17 Oct 2024 15:18:42	U	SMPL
P2403-10RE - 1	LOD-MDL-WATERRE	516	0.1256	-	245.1	PPB	17 Oct 2024 15:23:30	U	SMPL
P2403-12 - 1	MDL-WATER	263	0.0570	-	245.1	PPB	17 Oct 2024 15:25:45	U	SMPL
P4347-01LX5 - 1		114	0.0167	-	245.1	PPB	17 Oct 2024 15:28:01	U	SMPL
P4347-01A - 1		14140	3.8149	-	245.1	PPB	17 Oct 2024 15:30:17	U	SMPL
CCV37 - 1	CCV37	18969	5.1226	-	245.1	PPB	17 Oct 2024 15:32:33	U	SMPL
CCB37 - 1	CCB37	-219	-0.0735	-	245.1	PPB	17 Oct 2024 15:34:49	U	SMPL