

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132986

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/17/2024	13:15
2	CAL2	1	Water	NA	NA	20.2	7.00	10/17/2024	13:16
3	CAL3	1	Water	NA	NA	20.2	10.02	10/17/2024	13:20
4	ICV	1	Water	NA	NA	20.3	7.00	10/17/2024	13:22
5	CCV1	1	Water	NA	NA	20.3	2.01	10/17/2024	13:25
6	P4419-04	1	Solid	20.02	20	21.6	7.66	10/17/2024	13:35
7	P4419-04DUP	1	Solid	20.03	20	21.7	7.67	10/17/2024	13:36
8	P4425-02	1	Solid	20.02	20	21.5	7.25	10/17/2024	13:40
9	P4425-04	1	Solid	20.03	20	21.6	7.11	10/17/2024	13:42
10	P4425-06	1	Solid	20.04	20	21.7	7.00	10/17/2024	13:44
11	P4425-08	1	Solid	20.02	20	21.3	6.89	10/17/2024	13:47
12	P4425-10	1	Solid	20.01	20	21.3	6.91	10/17/2024	13:50
13	P4430-01	1	Solid	20.02	20	24.2	5.77	10/17/2024	14:00
14	CCV2	1	Water	NA	NA	20.3	12.02	10/17/2024	14:10