



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/21/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133000

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4442-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
P4442-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
P4442-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
P4442-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
P4442-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
P4442-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
P4442-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
P4442-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
P4442-13	PCB-GPC2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
P4442-14	PCB-GPC2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
P4443-01	OG-315-HR-502-COMP-29	11	1.17	8.40	9.57	8.9	92.0	
P4443-02	OG-315-HR-502-VOC-29	12	1.15	8.40	9.55	8.8	91.1	
P4443-03	OG-315-HR-502-57	13	1.17	8.60	9.77	9.00	91.0	
P4443-04	OG-315-HR-502-58	14	1.19	8.50	9.69	9.16	93.8	
P4443-06	OG-315-HR-502-COMP-30	15	1.18	8.41	9.59	8.4	85.9	
P4443-07	OG-315-HR-502-VOC-30	16	1.15	8.70	9.85	8.39	83.2	
P4443-08	OG-315-HR-502-59	17	1.15	8.81	9.96	9.33	92.8	
P4443-09	OG-315-HR-502-60	18	1.12	8.76	9.88	7.99	78.4	
P4450-01	PIPE-0	19	1.00	1.00	2.00	2.00	100.0	wipe sample
P4450-02	PIPE-3	20	1.00	1.00	2.00	2.00	100.0	wipe sample
P4450-03	PIPE-5	21	1.00	1.00	2.00	2.00	100.0	wipe sample
P4452-01	ETGI-285	22	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4454-01	34542	23	1.15	8.50	9.65	9.00	92.4	
P4454-02	34543	24	1.15	8.82	9.97	9.32	92.6	
P4454-03	34542-43	25	1.15	8.83	9.98	9.44	93.9	
P4455-01	SU-4-101824	26	1.12	8.49	9.61	8.85	91.0	
P4455-02	SU-4-101824-E2	27	1.12	8.84	9.96	9.07	89.9	
P4456-01	PAD-10182024	29	1.00	1.00	2.00	2.00	100.0	CONCRETE sample



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/21/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133000

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4457-01	KMA755I-END-1-1	30	1.00	1.00	2.00	2.00	100.0	pilc
P4457-02	KMA755I-END-1-2	31	1.00	1.00	2.00	2.00	100.0	pilc
P4457-03	KMA755I-END-2-1	32	1.00	1.00	2.00	2.00	100.0	pilc
P4457-04	KMA755I-END-2-2	33	1.00	1.00	2.00	2.00	100.0	pilc
P4457-05	BL270845LOK-END-1-1	34	1.00	1.00	2.00	2.00	100.0	pilc
P4457-06	BL270845LOK-END-1-2	35	1.00	1.00	2.00	2.00	100.0	pilc
P4457-07	BL270845LOK-END-2-1	36	1.00	1.00	2.00	2.00	100.0	pilc
P4457-08	BL270845LOK-END-2-1	37	1.00	1.00	2.00	2.00	100.0	pilc
P4457-09	BC260234LOK-END-1-1	38	1.00	1.00	2.00	2.00	100.0	pilc
P4457-10	BC260234LOK-END-1-2	39	1.00	1.00	2.00	2.00	100.0	pilc
P4457-11	BC260234LOK-END-2-1	40	1.00	1.00	2.00	2.00	100.0	pilc
P4457-12	BC260234LOK-END-2-1	41	1.00	1.00	2.00	2.00	100.0	pilc
P4457-13	SW633963-END-1-1	42	1.00	1.00	2.00	2.00	100.0	pilc
P4457-14	SW633963-END-1-2	43	1.00	1.00	2.00	2.00	100.0	pilc
P4457-15	SW633963-END-2-1	44	1.00	1.00	2.00	2.00	100.0	pilc
P4457-16	SW633963-END-2-2	45	1.00	1.00	2.00	2.00	100.0	pilc
P4458-01	280517	46	1.15	8.68	9.83	8.75	87.6	
P4459-01	60379	28	1.00	1.00	2.00	2.00	100.0	wipe sample
P4460-02	WB-303-TOP	47	1.19	8.52	9.71	5.58	51.5	
P4460-03	WB-303-BOT	48	1.15	8.81	9.96	8.21	80.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$