

6133007

=====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/18/2024 14:43

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.009	0.0	0.111	
ICB1	-0.005	0.0	0.008	
CCV1	1.010	0.0	0.111	
CCB1	-0.006	0.0	0.008	
PB164255BL	-0.005	0.0	0.008	
PB164255BS	1.015	0.0	0.111	
P4368-01	0.021	0.0	0.011	
P4368-02	0.046	0.0	0.013	
PB164257BL	-0.005	0.0	0.008	
PB164257BS	1.006	0.0	0.111	
P4368-07	0.021	0.0	0.011	
P4368-08	0.044	0.0	0.013	
CCV2	1.005	0.0	0.110	
CCB2	-0.005	0.0	0.008	

N 14
Mean 0.368
SD 0.4962
CV% 134.93

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 18 13:01:10 2024

Fri Oct 18 14:42:39 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0088	mg/l	10/18/2024 13:01:10	
0.05PPM	A	Phenolics-	P	0.059	mg/l	10/18/2024 13:01:11	
0.1PPM	A	Phenolics-	P	0.1076	mg/l	10/18/2024 13:01:12	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	10/18/2024 13:01:13	
0.50PPM	A	Phenolics-	P	0.4947	mg/l	10/18/2024 13:01:14	
1.0PPM	A	Phenolics-	P	0.9921	mg/l	10/18/2024 13:01:15	
2.0PPM	A	Phenolics-	P	2.0045	mg/l	10/18/2024 13:01:16	
ICV1	S	Phenolics-	P	1.0089	mg/l	10/18/2024 14:36:44	
ICB1	S	Phenolics-	P	-0.0045	mg/l	10/18/2024 14:36:45	
CCV1	S	Phenolics-	P	1.0097	mg/l	10/18/2024 14:36:46	
CCB1	S	Phenolics-	P	-0.0065	mg/l	10/18/2024 14:36:47	
PB164255BL	S	Phenolics-	P	-0.0054	mg/l	10/18/2024 14:36:48	
PB164255BS	S	Phenolics-	P	1.0147	mg/l	10/18/2024 14:36:49	
P4368-01	S	Phenolics-	P	0.0208	mg/l	10/18/2024 14:36:51	
P4368-02	S	Phenolics-	P	0.0457	mg/l	10/18/2024 14:36:53	
PB164257BL	S	Phenolics-	P	-0.0053	mg/l	10/18/2024 14:36:54	
PB164257BS	S	Phenolics-	P	1.0063	mg/l	10/18/2024 14:36:55	
P4368-07	S	Phenolics-	P	0.0208	mg/l	10/18/2024 14:42:34	
P4368-08	S	Phenolics-	P	0.0439	mg/l	10/18/2024 14:42:36	
CCV2	S	Phenolics-	P	1.0046	mg/l	10/18/2024 14:42:38	
CCB2	S	Phenolics-	P	-0.0054	mg/l	10/18/2024 14:42:39	

CHEMTECH CONSULTING GROUP INC
 284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

10/18/2024 13:06


Test Phenolics-

Accepted 10/18/2024 13:06

Factor 9.859
 Bias 0.008

Coeff. of det. 0.999897

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.008	-0.0088	0.0000	-
2	Phenol-2	0.014	0.0590	0.0500	18.6
3	Phenol-2	0.019	0.1076	0.1000	7.6
4	Phenol-2	0.034	0.2508	0.2500	0.3
5	Phenol-2	0.059	0.4947	0.5000	-1.1
6	Phenol-2	0.109	0.9921	1.0000	-0.8
7	Phenol-2	0.212	2.0045	2.0000	0.2

10/18/2024
 RM