

Lb133009

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

10/18/2024 15:41

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.004	0.0	0.110	
ICB1	-0.004	0.0	0.008	
CCV1	1.011	0.0	0.111	
CCB1	-0.003	0.0	0.008	
PB164256BL	0.001	0.0	0.009	
PB164256BS	1.013	0.0	0.111	
P4368-10	0.022	0.0	0.011	
P4368-11	0.049	0.0	0.013	
CCV2	1.012	0.0	0.111	
CCB2	-0.002	0.0	0.008	

N 10
Mean 0.410
SD 0.5164
CV% 125.87

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 18 15:11:30 2024

Fri Oct 18 15:19:21 2024

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0088	mg/l	10/18/2024 13:01:10	
0.05PPM	A	Phenolics-	P	0.059	mg/l	10/18/2024 13:01:11	
0.1PPM	A	Phenolics-	P	0.1076	mg/l	10/18/2024 13:01:12	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	10/18/2024 13:01:13	
0.50PPM	A	Phenolics-	P	0.4947	mg/l	10/18/2024 13:01:14	
1.0PPM	A	Phenolics-	P	0.9921	mg/l	10/18/2024 13:01:15	
2.0PPM	A	Phenolics-	P	2.0045	mg/l	10/18/2024 13:01:16	
ICV1	S	Phenolics-	P	1.004	mg/l	10/18/2024 15:11:30	
ICB1	S	Phenolics-	P	-0.0044	mg/l	10/18/2024 15:11:32	
CCV1	S	Phenolics-	P	1.0108	mg/l	10/18/2024 15:11:35	
CCB1	S	Phenolics-	P	-0.0026	mg/l	10/18/2024 15:11:36	
PB164256BL	S	Phenolics-	P	0.0011	mg/l	10/18/2024 15:11:38	
PB164256BS	S	Phenolics-	P	1.0131	mg/l	10/18/2024 15:11:40	
P4368-10	S	Phenolics-	P	0.0222	mg/l	10/18/2024 15:19:14	
P4368-11	S	Phenolics-	P	0.0486	mg/l	10/18/2024 15:19:15	
CCV2	S	Phenolics-	P	1.0118	mg/l	10/18/2024 15:19:18	
CCB2	S	Phenolics-	P	-0.0022	mg/l	10/18/2024 15:19:21	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

10/18/2024 13:06

Test Phenolics-


Accepted 10/18/2024 13:06

Factor 9.859

Bias 0.008

Coeff. of det. 0.999897

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	-0.0088	0.0000	-
2	Phenol-2	0.014	0.0590	0.0500	18.0
3	Phenol-2	0.019	0.1076	0.1000	7.6
4	Phenol-2	0.034	0.2508	0.2500	0.3
5	Phenol-2	0.059	0.4947	0.5000	-1.1
6	Phenol-2	0.109	0.9921	1.0000	-0.8
7	Phenol-2	0.212	2.0045	2.0000	0.2

10/18/2024

RM