

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133016

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/19/2024	12:20
2	CAL2	1	Water	NA	NA	20.2	6.98	10/19/2024	12:21
3	CAL3	1	Water	NA	NA	20.2	10.02	10/19/2024	12:22
4	ICV	1	Water	NA	NA	20.3	7.01	10/19/2024	12:25
5	CCV1	1	Water	NA	NA	20.3	2.01	10/19/2024	12:30
6	P4454-01	1	Solid	20.02	20	24.3	5.75	10/19/2024	12:37
7	P4454-01DUP	1	Solid	20.03	20	24.4	5.76	10/19/2024	12:38
8	P4454-02	1	Solid	20.05	20	24.2	6.60	10/19/2024	12:44
9	CCV2	1	Water	NA	NA	20.3	12.02	10/19/2024	12:50

WORKLIST(Hardcopy Internal Chain)

133016

WorkList Name : PH P4454 WorkList ID : 184610 Department : Wet-Chemistry Date : 10-19-2024 12:05:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4454-01	34542	Solid	pH	Cool 4 deg C	PSEG03	K51	10/18/2024	9045D
P4454-02	34543	Solid	pH	Cool 4 deg C	PSEG03	K51	10/18/2024	9045D

Date/Time 10/19/24 12:15
Raw Sample Received by: [Signature] 10001
Raw Sample Relinquished by: [Signature] 10001

Date/Time 10/19/24 14:00
Raw Sample Received by: [Signature] 10001
Raw Sample Relinquished by: [Signature] 10001