



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/23/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 10/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133030

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4416-01	A0BD7	1	1.15	8.81	9.96	8.6	84.6	
P4416-02	A0BD8	2	1.12	8.61	9.73	7.33	72.1	
P4416-03	A0BD8MS	3	1.12	8.61	9.73	7.33	72.1	
P4416-04	A0BD8MSD	4	1.12	8.61	9.73	7.33	72.1	
P4467-01	TP-1	5	1.15	8.84	9.99	9.17	90.7	
P4467-02	TP-1-EPH	6	1.17	8.60	9.77	8.89	89.8	
P4467-03	TP-1-VOC	7	1.15	8.57	9.72	8.87	90.1	
P4468-01	ETGI-331	8	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4468-03	ETGI-329	9	1.15	8.82	9.97	9.35	93.0	
P4468-05	ETGI-345	10	1.12	8.75	9.87	9.11	91.3	
P4469-01	16-PIPE	11	1.00	1.00	2.00	2.00	100.0	wipe sample
P4469-02	21-PIPE	12	1.00	1.00	2.00	2.00	100.0	wipe sample
P4469-03	32-PIPE	13	1.00	1.00	2.00	2.00	100.0	wipe sample
P4469-04	34-PIPE	14	1.00	1.00	2.00	2.00	100.0	wipe sample
P4470-01	CL-01-102124	15	1.12	8.75	9.87	9.39	94.5	
P4470-02	CL-01-102124-E2	16	1.12	8.76	9.88	9.36	94.1	
P4471-01	B-180-SB01	17	1.18	8.59	9.77	8.29	82.8	
P4471-02	B-180-SB02	18	1.16	8.80	9.96	5.49	49.2	
P4472-01	BP-F-28	19	1.16	8.46	9.62	8.75	89.7	
P4472-02	BP-F-28-VOC	20	1.18	8.45	9.63	8.92	91.6	
P4472-03	BP-F-28-EPH	21	1.15	8.64	9.79	8.94	90.2	
P4472-05	BP-F-6	22	1.18	8.58	9.76	9.00	91.1	
P4472-06	BP-F-6-VOC	23	1.15	8.83	9.98	9.31	92.4	
P4472-07	BP-F-6-EPH	24	1.16	8.83	9.99	9.35	92.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$