

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB133052

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP110348
Turbidity Calibration std, 0NTU	WP110346
Turbidity Calibration std, 20NTU	WP110341
Turbidity Calibration std, 0.5NTU	WP110345
Turbidity Calibration std, 40NTU	WP110340
Turbidity Calibration std, 1NTU	WP110344
Turbidity Calibration std, 5NTU	WP110343
Turbidity LOQ std, 1.0NTU	WP110347
Turbidity Calibration - CCV std, 10 NTU	WP110342
10 NTU Standard 500 ml	W3116

Intercept: 0.2418

Slope: 1.073

Regression: 0.9993

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.162	-0.07		10/22/2024	10:15
2	CAL2	1	1	1.260	0.95	-5.1	10/22/2024	10:18
3	CAL3	5	1	5.480	4.88	-2.4	10/22/2024	10:21
4	CAL4	10	1	11.900	10.87	8.7	10/22/2024	10:24
5	CAL5	20	1	20.800	19.16	-4.2	10/22/2024	10:27
6	CAL6	40	1	43.400	40.22	0.6	10/22/2024	10:30



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB133052

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.500	10.492	10/22/2024	10:33
2	ICB		1	0.168	-0.069	10/22/2024	10:36
3	CCV1	10	1	11.100	10.119	10/22/2024	10:39
4	CCB1		1	0.167	-0.070	10/22/2024	10:42
5	LB133052BL		1	0.165	-0.072	10/22/2024	10:45
6	P4368-07		1	0.828	0.546	10/22/2024	10:48
7	P4368-08		1	1.460	1.135	10/22/2024	10:51
8	CCV2	10	1	11.400	10.399	10/22/2024	10:54
9	CCB2		1	0.166	-0.071	10/22/2024	10:57