

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB133055

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP110346
Turbidity Calibration std, 20NTU	WP110341
Turbidity Calibration std, 40NTU	WP110340
Turbidity Calibration std, 1NTU	WP110344
Turbidity Calibration std, 5NTU	WP110343
Turbidity Calibration - CCV std, 10 NTU	WP110342
TURBIDITY LOD STD, 0.5NTU	WP110348
Turbidity LOQ std, 1.0NTU	WP110347
10 NTU Standard 500 ml	W3116

Intercept: 0.2418

Slope: 1.073

Regression: 0.9993

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.162	0.03		10/22/2024	10:15
2	CAL2	1	1	1.260	0.91	-9.5	10/22/2024	10:18
3	CAL3	5	1	5.480	4.85	-3.1	10/22/2024	10:21
4	CAL4	10	1	11.900	10.84	8.4	10/22/2024	10:24
5	CAL5	20	1	20.800	19.15	-4.3	10/22/2024	10:27
6	CAL6	40	1	43.400	40.24	0.6	10/22/2024	10:30



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB133055

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.500	10.492	10/22/2024	10:33
2	ICB		1	0.168	-0.069	10/22/2024	10:36
3	CCV1	10	1	11.100	10.119	10/22/2024	10:39
4	CCB1		1	0.165	-0.072	10/22/2024	10:45
5	LB133055BL		1	0.165	-0.072	10/22/2024	10:45
6	P4368-10		1	0.828	0.546	10/22/2024	10:48
7	P4368-11		1	1.460	1.135	10/22/2024	10:51
8	CCV2	10	1	11.400	10.399	10/22/2024	10:54
9	CCB2		1	0.166	-0.071	10/22/2024	10:57