

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133063

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/22/2024	16:45
2	CAL2	1	Water	NA	NA	20.1	7.01	10/22/2024	16:46
3	CAL3	1	Water	NA	NA	20.1	10.02	10/22/2024	16:48
4	ICV	1	Water	NA	NA	20.1	7.00	10/22/2024	16:50
5	CCV1	1	Water	NA	NA	20.1	2.02	10/22/2024	16:52
6	P4485-03	1	Water	NA	NA	24.2	6.27	10/22/2024	17:00
7	P4485-03DUP	1	Water	NA	NA	24.3	6.28	10/22/2024	17:01
8	P4488-03	1	Water	NA	NA	25.4	4.56	10/22/2024	17:15
9	P4488-05	1	Water	NA	NA	25.7	4.96	10/22/2024	17:22
10	CCV2	1	Water	NA	NA	20.2	12.02	10/22/2024	17:25

WORKLIST(Hardcopy Internal Chain)

VB 133063

WorkList Name : PH P4488 WorkList ID : 184671 Department : Wet-Chemistry Date : 10-22-2024 16:32:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4485-03	D20241001-02-03	Water	pH	Cool 4 deg C	PSEG03	K62	10/22/2024	9040C
P4488-03	HCC-1	Water	pH	Cool 4 deg C	PSEG03	K61	10/22/2024	9040C
P4488-05	HCC-2	Water	pH	Cool 4 deg C	PSEG03	K61	10/22/2024	9040C

Date/Time 10/22/24 16:40
Raw Sample Received by: TA WCI
Raw Sample Relinquished by: Carl Sw

Date/Time 10/22/24 17:50
Raw Sample Received by: Carl Sw
Raw Sample Relinquished by: JO WCI