



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/24/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 10/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:38
Out Date: 10/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133067

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4480-01	MYE490	1	1.16	8.56	9.72	9.13	93.1	
P4480-02	MYE491	2	1.16	8.61	9.77	9.51	97.0	
P4480-03	MYE492	3	1.14	8.52	9.66	9.02	92.5	
P4480-04	MYE493	4	1.15	8.81	9.96	9.15	90.8	
P4480-05	MYE494	5	1.13	8.62	9.75	9.33	95.1	
P4480-06	MYE494D	6	1.13	8.62	9.75	9.33	95.1	
P4480-07	MYE494S	7	1.13	8.62	9.75	9.33	95.1	
P4480-08	MYE495	8	1.14	8.38	9.52	9.21	96.3	
P4480-09	MYE496	9	1.15	8.46	9.61	9.15	94.6	
P4480-10	MYE497	10	1.17	8.48	9.65	9.28	95.6	
P4480-11	MYE498	11	1.15	8.62	9.77	9.59	97.9	
P4480-12	MYE499	12	1.17	8.52	9.69	9.42	96.8	
P4480-13	MYE4A0	13	1.16	8.39	9.55	9.33	97.4	
P4480-14	MYE4A1	14	1.17	8.46	9.63	9.24	95.4	
P4480-15	MYE4A2	15	1.17	8.44	9.61	9.17	94.8	
P4480-16	MYE4A3	16	1.16	8.76	9.92	9.53	95.5	
P4480-17	MYE4A4	17	1.15	8.66	9.81	9.42	95.5	
P4480-18	MYE4A5	18	1.17	8.71	9.88	9.57	96.4	
P4480-19	MYE4A6	19	1.17	8.53	9.7	9.42	96.7	
P4480-20	MYE4A7	20	1.16	8.37	9.53	9.31	97.4	
P4480-21	MYE4A8	21	1.15	8.55	9.7	9.52	97.9	
P4480-22	MYE4A9	22	1.15	8.76	9.91	9.69	97.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$